



Basic Light

User manual | British English

Some photographs and/or images in this manual may show different products from the one referred to.

However, all instructions are applicable, regardless of the product differences.

Total or partial reproduction of this manual is forbidden without written authorisation from Rehateam s.r.l.

Produced and published by Permobil

Edition: 0.1

Date: 2024-12-17

Order no: eng-UK

1	INTRODUCTION	5
2	FIRE RETARDANT PROPERTIES, SYMBOLS	6
2.1	FLAMMABILITY	6
2.2	SYMBOLS USED	6
3	PRODUCT LABELS	7
4	INTENDED PURPOSES AND CONDITIONS OF USE	8
5	PUTTING THE WHEELCHAIR INTO SERVICE AND REGULAR USE	10
5.1	GENERAL INFORMATION	10
5.2	TYRE PRESSURE	11
5.3	BRAKES	12
5.4	TYPE OF BRAKES	14
5.4.1	Push-to-lock brake	14
5.4.2	Brake with extended push lever	15
5.5	QUICK RELEASE REAR WHEELS	16
5.6	FOLDING AND UNFOLDING THE WHEELCHAIR	17
5.6.1	Unfolding	18
5.6.2	Folding	19
5.6.3	Backrest bar and linking system	19
5.7	FOOTREST	20
5.7.1	General information	20
5.7.2	Plastic dual footrest	22
5.7.3	Single flip-up plastic footrest	22
5.7.4	Single flip-up aluminium footrest	22
5.8	FRONT FRAME WITH REMOVABLE FOOTREST	23

5.8.1	General information	23
5.8.2	Front frame with removable footrest	24
5.9	HEIGHT-ADJUSTABLE FOOTRESTS	25
5.10	CALF STRAP	26
5.11	ADJUSTABLE BACKREST UPHOLSTERY	27
5.12	ARMRESTS	27
5.12.1	General information	27
5.12.2	Progeo Deskarmrest	29
5.12.3	Tubular armrests with standard support	30
5.12.4	Tubular armrests with steel supports	32
5.13	CLOTHING SIDE GUARD	34
5.14	FASTENING OF PARTS AND LIFETIME OF THE WHEELCHAIR	34
5.14.1	Parts fastening	34
5.14.2	Wheelchair lifetime	35
6	SAFETY	36
6.1	PREVENTING ACCIDENTS	37
6.1.1	Movements	37
6.1.2	Wheelchair speed	37
6.1.3	Use of the brakes	37
6.2	MOVEMENTS: GETTING IN AND OUT OF THE WHEELCHAIR	38
6.3	STARTING TO DRIVE YOUR WHEELCHAIR	40
6.4	OVERCOMING OBSTACLES: STEPS AND HEIGHT DIFFERENCES	41
6.4.1	Going up and down a step by yourself	42
6.4.2	Negotiating a step with an assistant	43
6.5	NEGOTIATING SLOPES OR INCLINES	44
6.6	ENSURING STABILITY	45

7	TRASPORT OF THE WHEELCHAIR	48
7.1	TRASPORT OF THE OCCUPIED WHEELCHAIR	49
7.2	TRANSPORT WITH AN ASSISTANT	49
7.3	LOADING THE WHEELCHAIR WITHOUT ASSISTANCE	50
8	DESCRIPTION OF PARTS	51
9	ADJUSTMENT	52
9.1	GENERAL INFORMATION	52
9.2	POSSIBLE ADJUSTMENTS	53
10	ACCESSORIES	54
10.1	GENERAL INFORMATION	54
10.2	NARROW PASSAGE TRANSIT WHEELS	54
10.3	ANTI-TIP WHEEL	55
10.4	ASSISTANT'S PUSH HANDLEBAR	56
10.5	TIPPING AID	57
10.6	CRUTCH SUPPORT	57
10.7	TRAY	58
10.8	PELVIC BELT AND HARNESS	59
10.9	HEADREST	61
10.10	LATERAL SUPPORT	62
10.11	REAR WHEELS	62
10.11.1	General information	62
10.11.2	Wheels Off-road	63
10.11.3	Wheels Fat-wheel	63
10.11.4	Wheels with single drive	63
10.11.5	22" and 24" wheels equipped with drum brakes	64
10.11.6	12" wheels (300mm)	65
10.12	HAND RIM	66

10.13	TYRES	67
10.14	SPOKE GUARDS	67
10.15	CASTORS	68
10.16	VANITY FLAP (seat/backrest bonding canvas)	69
10.17	REMOVABLE UTILITY BAG	69
10.18	FRAME PROTECTION PAD	70

11	MAINTENANCE	71
11.1	REPLACEMENT OF WORN PARTS	72
11.1.1	Repairing a puncture	72
11.1.2	Spare parts	72
11.2	CHECKING THE PARTS	72
11.3	DISINFECTION AND REUSE OF THE WHEELCHAIR	73
11.4	EXTENDED STORAGE	73
11.5	DISPOSAL/RECYCLING OF PARTS	74
12	TROUBLESHOOTING GUIDE	75
13	TECHNICAL SPECIFICATIONS	76
14	WARRANTY	77
15	WARRANTY CERTIFICATE	79

1 INTRODUCTION

At Rehateam, constant research into quality and creativity are the cornerstone of our business.

Rehateam s.r.l., a serious and reliable company, genuinely focused on providing total customer satisfaction.

We have become industry leaders by making excellence and service our top priorities.

We have become industry leaders by making excellence and service our top priorities, building all our models from carefully selected materials and providing multiple configurations to enable the utmost personalisation according to specific user needs.

We perform continuous, meticulous quality control and our testing procedures offer the highest possible quality combined with fast, reliable service.

We owe our success to our customers and all who have believed in us and helped us establish that quality is the supreme differentiator.

FOR THE DEALER

This manual **MUST** be provided to the occupant of this wheelchair.

BEFORE giving the wheelchair to the occupant, the dealer **MUST** carry out a general check of all its fittings and check its functionality.

After this check, the dealer should apply his stamp or that of the company he represents, his name and the date, on the **WARRANTY CERTIFICATE**, which must be kept by the user.

This certificate must be retained by the occupant.

This product is intended for disabled occupants with intact visual, cognitive and reading abilities.

If some or all of these abilities are limited or impaired, the presence of an assistant person with appropriate abilities is always required.

FOR THE OCCUPANT and/or CARER

BEFORE using this wheelchair, please **READ** all parts of this manual and keep it for future reference.

Check that the warranty certificate has been fully compiled by the dealer and keep it with care.

NOTE

This document is also available in PDF format for the visually impaired at www.rehateamprogeo.com which can also be accessed by scanning the QR code provided here.



2 FIRE RETARDANT PROPERTIES, SYMBOLS

2.1 FLAMMABILITY

This product has been tested in accordance with the standard EN12183. This standard also includes a flammability test.

The fire reaction of upholstered parts complies with classes: M2 and M3.

2.2 SYMBOLS USED



DANGER!

Indicates a potentially hazardous situation, which, if not avoided, can result in death or serious personal injury.



CAUTION!

This symbol indicates an action that is forbidden or not recommended.



WARNING!

Indicates a potentially hazardous situation, which, if not avoided, may cause minor personal injuries or property damage.
















IMPORTANT!

This symbol indicates important instructions or special information.

3 PRODUCT LABELS



 Progeo brand logo	 "UDI" code (Unique Device Identification)
 "MD" mark (Medical Device)	 Year/month of production
 "CE" mark	 Serial number
XXXXXXXXXX Product name	 Ambient usage temperature
 Crash test approved Suitable for transporting the occupied wheelchair in a vehicle	 Max Mamimum load
 Not suitable for transporting the occupied wheelchair in a vehicle	 Max Maximum slope in use
 Read the user's manual	 Manufacturer data

4 INTENDED PURPOSES AND CONDITIONS OF USE

This model is a self-propelled wheelchair which means that it **can be moved manually** by driving the hand rims on the rear wheels or by an assistant using the push handles.

If self-propulsion is achieved via the lower limbs, take care that the user's feet and/or legs do not collide with any part of the wheelchair.

It may be used for both rehabilitative and active purposes and is therefore suitable for daily use by both occupants with severe disabilities (with limited movement of the legs, arms or torso) and "active" occupants (capable of carrying out activities independently).

It is used to assure the independent mobility or assisted movement of an occupant with reduced or no use of one or more parts of the body.

Given the ample range of possible accessories and configurations, the wheelchair can be used safely both indoors (flat, gym, school, library, etc.) and outdoors (street, pavement, yard, etc.) with ambient temperatures between -30 and +60 °C.

It is possible that the surfaces of the wheelchair, which may come into direct contact with the skin of the occupant/assistant, will exceed a surface temperature of 41° (e.g. exposure to the sun), so take care before use.

If the occupant is affected by particularly severe disabilities or unable to move autonomously, the help of an assistant is always advisable.

We also discourage use over hilly, particularly soft, sandy or uneven ground, on slopes exceeding the recommended gradient and in acidic environments.



WARNING!

During daily use of the wheelchair, the occupant may come across situations such as rough or irregular terrains (gravel, cobblestones, holes on the ground, etc.) or slippery terrain (wet, sandy, dusty, oily hard surfaces, etc.).

In all such situations and, in general, in all cases where use of the wheelchair may be difficult or risky (for the wheelchair and its occupant), it is always advisable to move carefully and smoothly without sudden acceleration or hard braking.

When the difficulty or risk is high, the help of an assistant is always advisable.

The maximum slope that can be negotiated to prevent the wheelchair from tipping over backwards is 3° (6%).

All footrests require the occupant to wear socks or shoes, so do not rest bare feet on them.

The seating surface, in cloth or other material if available for this model, is not designed to be directly sat on by the occupant; it is therefore necessary to provide a cushion for the occupant to sit on and which must be ordered separately according to the occupant's specific needs.



WARNING!

Crushing, holes and clearances: the wheelchair meets the requirements of safety distances. If certain functional requirements do not comply with these distances, be careful when using it (toe and foot traps).



IMPORTANT!

In certain configurations, the wheelchair may **exceed 700 mm in width**.

In such cases, under certain circumstances, it may not be possible to use some or all of the available escape routes from a building.

If the wheelchair configuration causes the wheelchair width to exceed 700 mm, it may be difficult or impossible to travel on public transport, lifts and hoists.

In certain configurations, the wheelchair may **exceed 1200 mm in length**.

Therefore, it may be difficult, if not impossible, to use public transport and/or transport the occupied wheelchair (if provided).

Given its dimensional and structural characteristics, this model is ideal for use by both adolescents and adults adolescents and adults.

The maximum weight load for this model is: 125 kg.

5 PUTTING THE WHEELCHAIR INTO SERVICE AND REGULAR USE

5.1 GENERAL INFORMATION



WARNING!

It is important that all the wheelchair's functional parts are CHECKED ONLY BY QUALIFIED AND AUTHORISED PERSONNEL, both to assess its suitability and to provide the correct instructions for use to occupant and assistant where applicable.

If any part of the wheelchair is found to be defective, the occupant or assistant must consult the dealer to solve the problem.

Non-observance of the indications/instructions in this manual may lead to malfunctions and/or damage to the wheelchair and/or the occupant and/or third parties.

is not liable for any damage to the wheelchair and/or the occupant and/or third parties caused by or resulting from the non-observance of the indications/instructions.



CAUTION!

Use of the wheelchair requires some of its parts to be handled.

Instructions for the correct handling of the wheelchair's components are provided in the corresponding sections of this document.

Please follow these instructions to avoid injuries.



CAUTION!

The installation and use of electrification, mechanical and/or electrical drive devices (lifts, stair lifts) or any other accessory not provided by Rehateam s.r.l. at the time of delivery may result in stresses that were not foreseen during product testing and therefore the integrity and complete safety of the product is not guaranteed.

In case of installation of any additional accessory to the product delivered by Rehateam s.r.l., the latter must be guaranteed by your dealer, the accessory device manufacturer, and approved by Rehateam s.r.l..

5.2 TYRE PRESSURE

The tyres must always be inflated to the correct pressure to ensure that the wheelchair brakes and rolls properly.

Use your fingers to remove the plastic cap which covers the inflation valve on the rim of the rear wheel.

Use a compressor or pump equipped with a pressure gauge to bring the pressure up to the correct level.

Screw the plastic cap back on.




WARNING!

Check for correct tyre pressure on a weekly basis as indicated on the tyre (an intermediate value between min. and max. is recommended).

Proper tyre pressure makes the wheelchair considerably smoother and more controllable to manoeuvre.

Inadequate tyre pressure (especially if lower than prescribed) degrades the effectiveness of the parking brakes (the wheelchair may not remain stationary).





CAUTION!

Do not exceed the pressure indicated on the tyre because you may damage the tyre and the inner tube.

Always respect what indicated on the tyre.



IMPORTANT!

The front castors may also be of the "pneumatic" type; in this case, follow the same instructions.

Solid tyres do not require any inflation.

This type of tyre requires less maintenance and reduces puncture repair or replacement costs.

Compared to tubed or tubular tyres, they are not as smooth during normal use and absorb less vibration when rolling on uneven ground.

5.3 BRAKES



CAUTION!

The parking brakes are a safety device and their efficiency must be checked daily.

The pressure brake acts directly on the tyre by means of a lever, which is why braking efficiency depends on correct tyre inflation pressure.

We recommend that you check your tyre pressure regularly to ensure that it is always correct.

The maximum safe gradient for the parking brakes to work properly is 3° (6%).

The parking brake is a safety device, and you must not use it while the wheelchair is moving as this could cause the wheelchair to overturn with possible injury to the occupant.



IMPORTANT!

To check the effectiveness of the brakes, apply them as described below depending on the type of brake, and, with the person sitting in the wheelchair, try to gently push the wheelchair.

If the rear wheels do not turn or if they turn with great difficulty and still prevent unintentional movement of the wheelchair, the parking brakes are acting as required.

Check that both brakes, once applied, do not release accidentally.

Applying and releasing the parking brake may require more or less force depending on the type of brake chosen.

Occupants with reduced upper limb strength should use the extended lever.



WARNING!

Take care not to pinch/crush your fingers when applying and releasing the brakes, but also in any other situation where your hands come close to the brakes.

5.4 TYPE OF BRAKES

5.4.1 Push-to-lock brake

To engage the brake, push the lever forward until it locks.

To release the brake, pull the lever backward.

The brake automatically returns to its resting position.



5.4.2 Brake with extended push lever

To engage the brake, push the lever forward until it locks.

To release the brake, pull the lever backward.

The brake automatically returns to its resting position.

Extended lever: the lever may be folded forward to prevent the extended lever from obstructing the occupant's wheelchair transfer manoeuvre.

To do so, pull the lever first up and then push it forward.

To return it to the normal using position, turn the lever up and then push it down.



5.5 QUICK RELEASE REAR WHEELS

The rear quick release wheels enable you to reduce the wheelchair's bulk and save space, thus facilitating storage in the car or, if the wheelchair is equipped with the "narrow passage wheels" accessory, to reduce its width and length in order to access narrow spaces.

Check that the quick release pins on the rear wheels are correctly engaged without anyone sitting on the wheelchair.

The parking brakes must be released.

Raise the rear wheel by lifting the wheelchair slightly off the ground and grasping the backrest post handle.

Grasp the rear wheel hub with your hand (without pressing down on the release pin), taking care to avoid inserting your fingers between the spokes.

Push and pull the wheel (as indicated by the arrow) to check that it fits properly: the wheel must not come off.

If the wheel comes off during this check, the axle must be adjusted.

Adjustment should only be carried out by qualified and authorised personnel.




WARNING!

Checking the quick release axles of the rear wheels is one of the procedures that the occupant should perform periodically.

If the axles are incorrectly inserted and/or adjusted, during use the components could be damaged and the rear wheel could accidentally come off, causing the wheelchair to overbalance with possibly injury to the occupant.

To remove the wheel, first release the parking brakes.

Lift the wheelchair wheel slightly off the ground, using one hand to grasp the backrest knob to facilitate the operation.



Grasp the wheel by the spokes around the hub, press the release pin button with your thumb and, without releasing, extract the wheel fully.

To fit the wheel, first release the brake.

Slightly lift the wheelchair from the ground on one side by the backrest knob.

With your hand, take hold of the rear wheel's hub and, with your thumb, press the button of the quick release pin, keep it pressed and push the pin fully into its seat.

Release the axle quick release pin button (a click indicates correct engagement).



WARNING!

Always release the brake before removing and inserting the rear wheel. Always check the correct engagement of the quick release axles. The holes for the rear wheel positioning may pinch or crush your fingers. Take care whenever your hands are close to such holes.

5.6 FOLDING AND UNFOLDING THE WHEELCHAIR

This model has a crossbar system foldable seat.

5.6.1 Unfolding

Lean the palms of your hands on the seat posts, positions **(1)** and **(2)**, and push them down until the wheelchair opens fully.

After opening the wheelchair, carefully check that the seat posts are well inserted on the four supports **(3)**.

To make the operation easier, lift one rear wheel to reduce friction with the floor.

To do so, grasp the backrest and slightly raise the wheelchair.

Now, with your free hand on the seat post, position A or B, push the seat post down onto the supports S.

Check that the seat posts are well inserted on the four supports S.

Lower the footrest (this operation is not necessary if the wheelchair is equipped with an automatic closure footrest).



WARNING!

While opening the wheelchair, be careful not to squeeze your fingers between the seat post and the supports or between the seat post and the frame post



5.6.2 Folding

Remove the cushion or any other soft or rigid supports from the seat and the backrest.

Raise the footrest (this operation is not necessary if the wheelchair is equipped with an automatic closure footrest).

With both hands, pull up the seat fabric as shown in the image.



5.6.3 Backrest bar and linking system

The Maxi version, i.e. with seat widths of 48, 51 and 54 cm, is supplied with a folding backrest stiffening bar **(4)** and a footrest linking system **(5)**.

These two accessories are particularly useful for making the wheelchair more rigid.

After opening the wheelchair, press the rigidizer bar down and lower the footrests so that the two metal parts engage with each other.

Before closing the wheelchair, raise the footrests and push the stiffening bar upwards.



5.7 FOOTREST

5.7.1 General information



WARNING!

When transferring to and from the wheelchair, never keep your feet on the footrest; excessive weight on the footrest may result in it breaking with harm to the occupant.

After each operation, check that the footrest is always in the correct operating position.

Dual footrests are less resistant than the single version and they are therefore not suitable for occupants subject to strong spasms and/or contractions.



WARNING!

Do not use the wheelchair with bare feet to avoid possible abrasions and/or burns to the feet.

Make sure that the surfaces of the footrest are in good condition as potentially dangerous sharp areas may be created if they come into contact with other surfaces (ground, walls, furniture, etc.).

Depending on the occupant's needs, the wheelchair can be provided with either of the above footrest versions.

In the order form, according to the type of footrest, you can choose its position (internal, external or intermediate)

Single footrests are more rigid than the dual versions and they are recommended for occupants subject to frequent spasms or who apply a lot of force to the footrest.

All footrests, except those with automatic closure, when raised, enable the occupants to rest their feet on the ground more easily, especially when transferring to and from the wheelchair.

Flipping up the footrest is a very easy operation you can perform with one hand by pulling the footrest plate upward.

To put the footrest back to the operating position, just push it down and, for a single footrest, make sure it locks in the support arm as indicated for each type of footrest.


Other information on footrests

To comply with the rule UNI EN 12183 paragraph 6.1, when dual footrests are fitted, it is necessary to maintain the correct clearance between the two footrest plates when lowered:

The distance between the two footrest plates “C” must not exceed 35 mm for wheelchairs used by adults (1).

The distance between the two footrest plates (C) must not exceed 25 mm for wheelchairs used by children (2).

The image on the side shows the correct dual footrest size (“B”) with respect to seat width (“A”):




5.7.2 Plastic dual footrest

Make sure both footrests are fully lowered before using the wheelchair.



5.7.3 Single flip-up plastic footrest

Make sure that the locking pin (1) is properly inserted in the recess of the support (2).



5.7.4 Single flip-up aluminium footrest

Make sure that the locking pin (1) is properly inserted in the recess of the support (2).



5.8 FRONT FRAME WITH REMOVABLE FOOTREST

5.8.1 General information

One of the features of this model is the particular footrest that you can rotate out of the way or remove completely.

These functions reduce the bulk and the weight of the wheelchair and help the occupant to get closer to objects such as beds and tables.

It is also convenient for amputees who have no use for the footrest.

5.8.2 Front frame with removable footrest

Make sure that the locking pin (1) is properly inserted in the recess of the support (2).

OPERATING POSITION

Check that the footrest frame is securely locked to its support.

Take hold of the frame with one hand (1) and try turning it to the right and left. If it is securely locked, the footrest frame will not rotate, but you should notice just a little play.

ROTATION

To swing the footrest away, push the lever (2), then, turn the footrest away outwards (3).

To close the footrest again, turn it in the opposite direction 4 until a click indicates that it is locked.

REMOVING/INSERTING

Push the lever (2) and pull the footrest frame up to release it (5).

To insert the footrest, align the two holes of the winglets (6) of the footrest frame with the front frame pins (7).


Insert the footrest frame and slit it in while keeping it partly swung away.

Then turn the footrest inwards to close it in the operating position (1).

Make sure it is securely locked following the previously outlined instructions.

IMPORTANT!

If the wheelchair is equipped with a single footrest, it must always be raised before swinging the footrest away.



5.9 HEIGHT-ADJUSTABLE FOOTRESTS

This type of frame allows to change the angle of the footplate's frame and thus of the posture of the lower limbs.

The height-adjustable footrests can also rotate and are removable (see instructions in chapter 5.8.2 *Front frame with removable footrest*).

To raise the footrest (change its inclination), simply hold it at the bottom **(1)** and rotate it upwards to the desired position.

The locking unit **(2)** will maintain the desired inclination.

To lower the footrest, hold it at the bottom, push back and hold the pin of locking unit **(2)** in this position and guide the footrest downwards.

Once the desired inclination has been reached, release pin **(2)**.



WARNING!

Before setting the locking unit free by releasing the pin **(2)**, always take hold of the lower side of the footrest frame, otherwise it will drop suddenly creating a possible risk to the occupant.



5.10 CALF STRAP


In some cases, depending on the diameter of the front wheel and the width of the wheelchair, the front wheel may come into contact with the heel of the occupant's foot when turning.

To prevent this contact and prevent the wheelchair from tipping over, the wheelchair is equipped with a heel strap that must be adjusted so that the occupant's feet never come into contact with the front wheels.



IMPORTANT!

You can easily adjust the calf strap using the Velcro straps provided.



5.11 ADJUSTABLE BACKREST UPHOLSTERY

The backrest can be adjusted to the occupant's requirements by tensioning or slackening the two special Velcro straps fitted inside the backrest itself.

To carry out the adjustment, lift or remove the upholstery, tighten, or slacken the Velcro straps as required and replace the upholstery.



CAUTION!

The combination of the inclination of the seat with the inclination of the backrest posts and the inclination of a postural backrest with respect to the posts, could result in an angle greater than 25° with respect to the vertical.

In that case, the wheelchair must have a headrest or arrangement to secure it as intended for this model.



IMPORTANT!


If the wheelchair is provided with a postural backrest, follow the instructions of the postural system.

5.12 ARMRESTS

5.12.1 General information

This model can be equipped with armrests.

The armrests are particularly useful accessories for people with reduced hand or arm mobility and for occupants with limited upper body movement who require better stability once seated in the wheelchair.





CAUTION!

Do not use the armrest as a hand hold to raise the wheelchair; this may cause the armrests to come away from their supports, damaging the wheelchair and harming the occupant.

Do not use the armrest for support when transferring to or from the wheelchair.

On the contrary, we recommend removing the armrest on the transfer side to facilitate the manoeuvre.



WARNING!

Armrest pads may contain phthalates that are potentially hazardous to health if ingested.

Avoid scratching the padding and putting any residue on your fingers into your mouth or, worse still, nibbling or sucking on the padding itself (as may occur with children).



WARNING!

Always pay attention to the positioning of hands and arms on the armrest to avoid accidents.

5.12.2 Progeo Deskarmrest

This type of armrest is very practical and can be removed completely, making it much easier to move the wheelchair and transport it.

The armrests also comprise a straight clothing guard.

OPERATING POSITION

To check that the armrest is in the correct operating position, take hold of the armrest pad **(1)** with one hand and try to pull it up.

If correctly secured, the armrest will not come off the support **(2)** and the pad **(1)** itself will not move.

REMOVING/INSERTING

To remove the armrest, press the lever **(3)** then take hold of the armrest pad and pull it off.

To insert the armrest, simply slide its part **(4)** into the support **(2)**.

Then lift the lever **(3)** to lock the armrest.

A click will indicate it has locked properly.

Check that the armrest is securely locked as described above.

ADJUSTMENT

The occupant can also easily adjust armrest height.

Push the button **(5)** forward, hold it and then raise or lower the armrest until it reaches the new height.

Release the button **(5)** and guide the armrest until it locks in one of the holes **(6)** in the tube.





WARNING!

Always check the armrests are securely locked in their using position.

If this instruction is ignored, the armrest may disengage and the occupant may fall from the wheelchair.

Be careful not to pinch or crush your fingers when fully lowering the armrest pad.

5.12.3 Tubular armrests with standard support

You can swing this type of armrest away on the side and remove it to make transfer easier.

Clothing guards are provided with this type of armrest.

OPERATING POSITION

To check that the armrest is in the correct operating position **(1)**, take hold of the armrest **(2)** and try to move it left and right.

If it is properly secured, the armrest will not turn but will have just a little play.

The armrest is correctly secured when the two flat faces of the support **(3)** engage between the two walls of the support **(4)**.

ROTATION


To facilitate wheelchair transfers, the armrest can be rotated: lift it about 2 cm **(5)** and swing it away until it locks in the opposite position, if desired **(6)**.

To return the armrest to the operating position, swing it in until it locks (if it was secured in its opposite position, first raise it by approximately 2 cm).

REMOVING/INSERTING

To remove the armrest, just pull it upward.

To insert the armrest; insert the support pin **(7)** into the hole in the support **(3)** and turn the armrest until it locks as described above.





WARNING!

This type of armrest can always be vertically extracted from its support.

Always check that the armrests are secured when they are in the operating position **(1)** (facing forwards and parallel with the sides of the wheelchair).

In all other positions (except when locked in their rear-facing position), they are free to turn and for this reason, you should always ensure that they are correctly positioned before using the wheelchair.

Failure to follow this instruction may cause the wheelchair or the occupant to overbalance.



IMPORTANT!

The supports are precisely machined so that inserting and removing the armrest is possible only if the movement is perfectly vertical.

5.12.4 Tubular armrests with steel supports

You can swing this type of armrest away on the side and remove it to make transfer easier.

Clothing guards are with this type of armrest.

OPERATING POSITION

To check that the armrest is in the correct operating position **(1)**, take hold of the armrest **(2)** and try to move it left and right.

If it is properly secured, the armrest will not turn but will have just a little play.

The armrest is correctly secured when the two flat faces of the support **(3)** engage between the two walls of the support **(4)**.

The correctly attached steel support does not allow unintentional removal of the armrest.

Check this by pulling the armrest upwards; it should stop after about 1cm.

ROTATION

To facilitate wheelchair transfers, the armrest can be rotated: lift it by approximately 1 cm and swing it sideways (part **(4)** of the support moves away from part **(3)** but it will not come off).

If you wish, you can swing the armrest away until it locks in its opposite position (part **(4)** locks into part **(3)**).

To return the armrest to operating position, repeat the steps described above.

During rotation, the two supports **(3)** and **(4)** remain raised about 1 cm, but do not come apart.

REMOVING/INSERTING

To remove the armrest, with one finger, move the pin **(8)** to align with the hole and then pull the armrest up.

To insert the armrest; insert the part **(7)** of the support into the hole of part **(3)** and turn the armrest until it locks as described above.





WARNING!

Always check that the armrests are secured when they are in the operating position **(1)** (facing forwards and parallel with the sides of the wheelchair).

In all other positions (except when locked in their rear-facing position), they are free to turn and for this reason, you should always ensure that they are correctly positioned before using the wheelchair.

Failure to follow this instruction may cause the wheelchair or the occupant to overbalance.



IMPORTANT!

The supports are precisely machined so that inserting and removing the armrest is possible only if the movement is perfectly vertical.

5.13 CLOTHING SIDE GUARD

The lateral clothing guards are in plastic and of the “mudguard” type. These are originally fixed with their edges approximately 5 mm above the tyre to prevent occupants from inserting their fingers between tyre and mudguard itself.



CAUTION!

Make sure that the surface of the clothing guards is in good condition as they may be sharp and could injure the user.



IMPORTANT!

Do not use the clothing guard for support when transferring to or from the wheelchair.

5.14 FASTENING OF PARTS AND LIFETIME OF THE WHEELCHAIR

5.14.1 Parts fastening

When delivered to the occupant, the wheelchair is checked by the retailer who thus ensures it is in perfect condition.

However, daily use of the wheelchair may cause bolts and nuts to loosen.





WARNING!

We recommend checking that all hardware is correctly tightened at least once a month. If necessary, request assistance from authorised Progeo wheelchair maintenance personnel.

At each maintenance cycle, ensure that all nuts, bolts, and screws are correctly tightened.

Failure to observe this warning could result in damage to the occupant, third parties or the vehicle, for which Rehateam s.r.l. shall not be liable.

5.14.2 Wheelchair lifetime

Progeo wheelchairs are used on a daily basis and they are consequently subject to continuous stress that causes inevitable wear and tear of the parts.

With this in mind, Progeo wheelchairs have a lifetime of 5 years provided they receive thorough periodic maintenance.

The lifetime will be considerably extended if the wheelchair is used only indoors or not on daily basis.

6 SAFETY



CAUTION!

The Progeo wheelchair is a medical device and therefore it **MUST NOT BE LENT NOR GIVEN TO OTHER USERS**, even for a short period.

It is forbidden to make any modifications to the wheelchair without our approval.

Modifying the wheelchair could create stability and strength problems with possible consequences for the user.



WARNING!

Rehateam s.r.l. will not be liable for failure to comply with the instructions or recommendations set out in this manual and any such activity will result in the immediate cancellation of the manufacturer's warranty.

BURN HAZARD! Wheelchair components may heat up if they are exposed to strong sunlight.



IMPORTANT!

Notify the manufacturer in the event of serious accidents in relation to the device.

Manufacturer: Rehateam s.r.l., Vicolo Negrelli 5, 31038 Castagnole di Paese TV, Italia, tel. 0039 0422 484657, mail: info_it@permobil.com

6.1 PREVENTING ACCIDENTS

6.1.1 Movements

Avoid sudden, jerky movements when using the wheelchair as this could cause it to overbalance.

If your path is obstructed by obstacles, avoid sudden movements (such as abrupt braking). There is increased risk of skidding on irregular or damp ground.

Passing over obstacles as steps and ramps should be done carefully (ask an assistant for help).

For greater safety, an assistant should help to push the wheelchair when moving across inclined or potentially dangerous terrain.

6.1.2 Wheelchair speed

Always adjust your speed in relation to the type of terrain and conditions.

As general rule, we advise a constant, regular speed avoiding sudden accelerations or changes in direction.

6.1.3 Use of the brakes

The parking brakes have been designed to hold the wheelchair stable when it is stationary.



CAUTION!

In the interest of safety, you should not use the parking brakes while the wheelchair is in motion.



WARNING!

When using the brakes, great care should be taken to avoid injury to the hands or fingers.

6.2 MOVEMENTS: GETTING IN AND OUT OF THE WHEELCHAIR

Wheelchair transfers must always be performed with extreme care and caution, even by experienced occupants, and only after receiving instruction from fully trained specialists. If wheelchair transfers are difficult, an assistant should be lend help. All these actions should in any case be performed taking account of the degree of your disability and your consequent level of autonomy.

The general rules to respect while getting in or out of the wheelchair safely are:

Ensure the wheelchair is parked on a solid, flat, or regular surface.

Do not get in or out of the wheelchair while on hills or uneven ground which could render the wheelchair unstable and cause the occupant to fall and/or the wheelchair to overbalance. The parking brakes should be engaged.

Move your feet off the footrest to get down. Move your feet close to the footrest to go up.


If the wheelchair is equipped with flip-up footrest (single or dual), raise it to facilitate the transfer.

(1) Lean on the wheelchair and, where possible, on a stable object nearby. Use the force of your arms to raise and move your body.

(2) If the occupant is not able to perform this movement alone, or if the movement is to be performed on an uneven surface, an assistant should provide help.

Assistants who help with wheelchair transfers must always do so with extreme caution.

Such operations must be performed only after receiving instruction from fully trained specialists.



6.3 STARTING TO DRIVE YOUR WHEELCHAIR

FINDING THE CENTRE OF GRAVITY

Your ability to control and safely tip the wheelchair mainly depends on your seating position and the location of the centre of gravity with respect to the rear wheels.

The correct position of the rear wheels depends on different factors such as the occupant's weight, type of disability and the occupant's driving skills.

The manufacturer delivers the wheelchair and the position of the rear wheels (configuration) as requested in the order form.

To find the point at which the chair will tip back easily and, consequently, to find the safe manoeuvring range, proceed as described below.

Starting from your seated wheelchair position:

Move forward by firmly grasping the rear wheel rims.

Then move the hand rims slightly backwards and, while moving your weight backwards, push the hand rims forward with a jerk to make the front of the wheelchair rise

At this point, by moving your body gently backwards and forwards, while moving the rear wheel backwards and forwards using the hand rims, you will be able to locate the centre of gravity and the different points at which you can achieve controlled tipping and risk dangerous overbalance of the wheelchair.



WARNING!

Finding the point at which the chair will tip back safely requires the presence of an assistant who must stand behind the wheelchair ready to prevent it from overbalancing.





WARNING!

If you hang a backpack, a bag or any other object on the push handlebars or any rear area of the wheelchair, the resulting uneven weight distribution can significantly affect the risk of overbalancing.

TO PREVENT THE WHEELCHAIR FROM OVERBALANCING DURING NORMAL USE, WE RECOMMEND FITTING AN ANTITIPPING SYSTEM.

6.4 OVERCOMING OBSTACLES: STEPS AND HEIGHT DIFFERENCES



WARNING!

Always take the utmost care when dealing with the situations described below.

Before negotiating a step, or overcoming an obstacle in general, on your own (i.e. without the presence of an assistant), you must have good control of your wheelchair and know how to Perform controlled tipping (wheelies), keeping your balance on the rear wheels.

Before negotiating a step, or overcoming an obstacle in general, make sure the anti-tipping system is in the neutral position.



DANGER!

Never attempt to negotiate steps or other obstacles that are over 20 cm in height.

When faced with particularly challenging obstacles, always seek assistance to ensure that such manoeuvres are performed with ease and safely.

6.4.1 Going up and down a step by yourself

GOING DOWN

Take the wheelchair right up to the obstacle so that the front wheels are as close as possible to it.

Raise the front castors by tilting the wheelchair while maintaining balance.

Slide the rear wheels very carefully over the edge.

Always keep your hands firmly on the hand rims during this step.

Once the rear wheels have descended the obstacle, tilt the wheelchair forward until the front castors touch the ground again.



GOING UP

Bring the wheelchair right up to the leading edge of the obstacle.

Raise the front wheels and tip the chair back while maintaining balance.

Using the hand rims, carefully move the rear wheels forwards until they touch the edge of the step.

Tip the wheelchair forward until the front castors rest on the upper surface of the obstacle.

Lean forward in the chair so that the rear of the chair is lighter.

Grasp the hand rims tightly and move the rear wheels forwards that will cause the rear wheel to slide over the edge of the step lifting the rear of the wheelchair off the ground resulting in the overcoming of the obstacle.



6.4.2 Negotiating a step with an assistant

GOING DOWN

Take the wheelchair right up to the obstacle so that the front wheels are as close as possible to it.


Grip the rear handles tightly and push down to raise the front of the wheelchair.

Keep the wheelchair in this position and accompany it down the step.

The occupant seated in the chair can help the assistant by using the hand rims.

The occupant seated in the chair can help the assistant by using the hand rims.

At the end of the descent, tip the wheelchair forward so that the front castors touch the ground again.




GOING UP

Move backwards towards the step so that the rear wheels of the wheelchair touch it.

Grip the rear handles tightly and push down to raise the front of the wheelchair.

Pull hard while keeping the wheelchair tilted (with the front wheels raised) to prevent the occupant from slipping out of the chair) until the rear wheels are over the step.

Keeping the wheelchair tilted, move it away from the step enough to enable the front castors to touch the same level ground.



GOING UP AND DOWN A SERIES OF STEPS



DANGER!

In the case of stairs with more than one step, two assistants are always required.


The first assistant stands behind the wheelchair and holds the push handles.

The second assistant grasps a fixed part of the front frame, guiding the wheelchair manoeuvre from the front.



IMPORTANT!

To make raising the front of wheelchair easier for the assistant, the wheelchair should be fitted with the “tipping aid” accessory.



6.5 NEGOTIATING SLOPES OR INCLINES

GOING DOWNHILL

To go downhill safely, the occupant should keep constant, controlled speed, and direction over the entire slope.

Approach the slope at a moderate speed.

Keep your weight backwards in the chair to avoid slipping out.

Hold on to the hand rims on the rear wheels and allow them to slide slowly through your hands.

The occupant should be able to stop the wheelchair at any time simply by blocking the motion of the hand rims.



GOING UPHILL

Driving up any kind of slope should be done with extreme care.

The occupant must move at speed by generating force on the rear wheel while at the same time maintaining control over the wheelchair.

Lean forwards and keep your weight forward in the chair to prevent the wheelchair from tipping over backwards.

Grasp the hand rims tightly.

Drive up the ramp by pushing the wheelchair quickly and forcefully on both hand rims.

Do not jerk, but move steadily and as smoothly as possible.

Do not swing your upper body.



IMPORTANT!

We recommend that occupants who have not yet completely mastered safe use of the wheelchair should have the anti-tip system fitted, which is useful for preventing the wheelchair from tipping over backwards.



CAUTION!

The maximum gradient that can be attempted in complete safety is 3° (6%).

6.6 ENSURING STABILITY

You will encounter situations which will require you to lean out of the wheelchair.

These apparently simple movements could, if not performed with care, result in a loss of stability and possible overbalancing of the wheelchair.

To achieve maximum occupant control of the wheelchair, we illustrate some situations which may emerge during normal daily use of the wheelchair. Please pay particular attention to the following to maintain balance and stability at all times.



WARNING!

Do not lean too far forward; you may fall out of the wheelchair.

Do not move forward by sliding your pelvis across the seat cushion to reach objects that are too far away, you could fall out of the wheelchair.

Do not lean too far back over the backrest, the wheelchair could overbalance.

Do not lean too far out to the side, the wheelchair could overbalance

LEANING FORWARDS


Ensure that the front castors are pointing forwards.

To do this, move the wheelchair forwards and then backwards.

Apply the parking brakes to prevent the chair from moving suddenly during the movement.

Lean forward in such a way that your upper body does not move beyond the front castors at any time (moving the weight too far forward may cause the wheelchair to tip up onto the front wheels and overbalance, possibly causing damage to the chair and harming the occupant).

To ensure greater stability, the occupant should hold on to the wheelchair with his or her free hand. Do not apply the parking brakes.



LEANING BACKWARDS

Ensure that the front castors are pointing forwards.

To do this, move the wheelchair forwards and then backwards.

Do not apply the parking brakes.


Lean backwards without changing your sitting position.



STRETCHING SIDEWAYS

Do not move your upper body beyond the rear wheel as the wheelchair could overbalance.

For greater safety and stability, hold on to the wheelchair with your free hand



7 TRANSPORT OF THE WHEELCHAIR

There is no single best way to load the wheelchair into a vehicle.

Factors include the type and degree of the occupant's disability (his/her ability to control his/her upper body, arm, and hand movements), physical strength (the operation may be too difficult for an older person or a child) and the type of vehicle that will be used.

It should be clear that all these factors are too many to indicate a single precise procedure to fit all cases.

For this reason, we will limit ourselves to general advice.



WARNING!

Perform all car loading activities with extreme care and only after receiving instruction from specialist service personnel at our authorised dealers.

If these procedures seem to be unsafe or difficult, we recommend asking an assistant for help.



CAUTION!

Never transport an occupant sitting on the wheelchair in a vehicle unless the wheelchair is certified for the purpose according to the requirements of the regulations.

If the occupant must be transported while seated on the wheelchair, remember that the wheelchairs are not supplied with seatbelts. Any safety belts for vehicle transport must be installed only by companies that are specialised in vehicle transformations.

7.1 TRANSPORT OF THE OCCUPIED WHEELCHAIR

Always make sure the wheelchair is suitable to transporting the user in a vehicle.

The symbols on the product label indicate if the wheelchair is or not crash tested and, accordingly, if it can be used for the transport of the user on a motor vehicle.

WARNING! Only wheelchairs marked with this symbol are judged suitable for use with the passenger on board the wheelchair.



7.2 TRANSPORT WITH AN ASSISTANT

In cases where the physical limitations of the occupant prevent him/her from moving and loading the wheelchair into the car on his/her own, the help of an assistant will be required.

The lightness and foldability of this model are very important because they enable the wheelchair to be moved easily even by persons with limited physical strength.

Remove all removable parts.

Release the brakes and remove the rear wheels.

Fold the wheelchair (seat and/or backrest).

Take hold of the chair with two hands on the two fixed parts of the frame to balance the weight of the wheelchair and make it easier to lift.

Loading the wheelchair without assistance.

The following actions are intended for people with good control of their upper body, arms, and hands as well as sufficient strength to perform the movements in complete safety.

Thanks to its reduced encumbrance, the wheelchair can be loaded in the boot of the car or between the front and rear seats.



images are purely indicative

In larger means of transport, you can load the wheelchair without reducing its encumbrance.

7.3 LOADING THE WHEELCHAIR WITHOUT ASSISTANCE

The following actions are intended for people with good control of their upper body, arms, and hands as well as sufficient strength to perform the movements in complete safety. Remove all removable parts.

Open the door on the side where you will get into the car from (e.g. the driver's door), then move the chair as close as possible to the seat.

Engage the brakes and get into the car.

Once you are inside the car, release the wheelchair brakes and pull out the rear wheels.

Fold the wheelchair (seat and/or backrest).

Take hold of the chair with two hands on the two fixed parts of the frame to balance the weight of the wheelchair and make it easier to lift.


Loading the wheelchair without assistance.

The following actions are intended for people with good control of their upper body, arms, and hands as well as sufficient strength to perform the movements in complete safety.



images are purely indicative

8 DESCRIPTION OF PARTS



1. Backrest post with handle
2. Parking brake
3. Tyre
4. Hand rim
5. Rear wheel
6. Rear frame
7. Rear wheel adjustment plate
8. Quick release axle
9. Fork support
10. Fork
11. Tight passage transit wheel and anti-tip wheel unit
12. Backrest upholstery + Velcro straps for adjustment
13. Armrest
14. Seat canvas
15. Crossbar
16. Footplate frame
17. Front frame
18. Footrest post
19. Heel strap
20. Footrest
21. Castor

9 ADJUSTMENT

9.1 GENERAL INFORMATION

The Progeo wheelchair is a medical device manufactured according to a patient-specific data sheet compiled by qualified personnel.

Rehateam s.r.l. manufactures, tests and delivers the wheelchair to the dealer with the measurements and characteristics required in the order data sheet.



CAUTION!

We strongly advise occupants against lending the wheelchair to other users even for brief periods.

Lending your wheelchair to others could cause it to function unsafely and tip over with potentially serious consequences both for the wheelchair and its occupant.

The wheelchair has a configuration and dimensions that are specially designed for the occupant who purchases it, and the characteristics of stability, manoeuvrability and resistance are guaranteed only for that occupant.

It is forbidden to carry out any modifications to the original design, even if they are possible.



CAUTION!

Adjustments may be carried out exclusively by qualified and authorised Rehateam s.r.l. personnel.

Any adjustments and/or modifications carried out by non-authorised persons will immediately void the product's warranty and relieve Rehateam s.r.l. from any liability concerning any malfunction and/or damage due to such adjustments/modifications.

Always contact Rehateam s.r.l. and its technicians for any non-standard requirements or modifications to enable them to assess such modifications and verify that they will not compromise the normal and safe use of the wheelchair.

Any modification of the original parameters and configuration could seriously compromise the safe operation of the wheelchair causing damage the wheelchair itself and harm to its occupant.

After every adjustment made to the wheelchair, check carefully that all parts are correctly fixed. Check that all screws and nuts are tightened and that all moving parts are functioning correctly.

After any adjustment, always test the wheelchair before normal use, possibly assisted by an assistant or a technician.

The possible adjustments for this model are listed below

The instructions to properly carry out the adjustments are included in the corresponding service manual that is available on the web site www.rehateamprogeo.com

9.2 POSSIBLE ADJUSTMENTS

- Rear height
- Front height
- Trim (centre of gravity)
- Fork angle
- Backrest height
- Backrest inclination
- Footrest distance
- Footrest position
- Footrest inclination
- Brakes
- Seat depth

10 ACCESSORIES

10.1 GENERAL INFORMATION

Each model can be equipped with various accessories that should be evaluated according to the needs of the occupant and assistant.

Some of the images in this manual may show a wheelchair model different from the reference one.

However, all instructions are applicable, regardless of the model differences.

10.2 NARROW PASSAGE TRANSIT WHEELS

This accessory becomes necessary when the dimensions of the wheelchair impede getting into or passing through narrow rooms such as a door or a lift.

To use this accessory, you should remove the rear wheels.

When using this accessory, you should turn the anti-tip wheel upward (see “*anti-tip wheel*”).



WARNING!

Remember that while using the transit wheels (with the rear wheels removed), the parking brakes will not work at all. The stability of the wheelchair resting on the wheels for narrow passages is reduced, therefore pay great attention and seek help from an assistant.



10.3 ANTI-TIP WHEEL

This accessory is designed to prevent the wheelchair from tipping backwards. To be efficient, it is fixed at approximately 2-3 cm from the ground.

You can have the anti-tip wheel can be fitted on one or, for more greater safety, on both sides of the wheelchair.

OPERATING POSITION

To check that the armrest is in the correct operating position **(1)**, take hold of the the post with one hand and try to pull it: the post must not come off.

ROTATION

To turn the anti-tip wheel up **(3)** or down (operating position), or press the pin **(2)** and rotate the anti-tip wheel.

REMOVING/INSERTING

To remove the anti-tip wheel, press the pin **(2)** and pull out the post.

To insert the anti-tip wheel, slide it inside the post **(4)**, turn it up or down as needed, and then make sure that the pin **(2)** slots into hole **(5)**.

NEUTRAL POSITION

Set the anti-tip wheel in neutral position **(3)** if necessary, for example, when going up steps.




WARNING!

Verify the efficiency of the system with an assistant.

The front of the wheelchair must be raised until the anti-tip wheel touches the ground and prevents the wheelchair from tipping.

Take great care when performing this operation.

Make sure it is correctly fitted (in the use position) before using the wheelchair.





IMPORTANT!

When overcoming obstacles or steps, the anti-tip wheel must be placed in neutral position.

10.4 ASSISTANT'S PUSH HANDLEBAR

You can assemble these only on a backrest that has posts without handles.

You can adjust their height to offer the assistant a better handhold and pushing position.


The assistant or the occupant can easily adjust and remove them.

To adjust their height, unscrew the lever **(1)**, slide the post up or down to the desired height, position the handle backwards and finally screw the lever **(1)** tightly.

The handle must not turn unintentionally.

To remove the push handle, unscrew the lever **(1)** and slide the post off the support.

To fit the push handle on the wheelchair, insert the post into the support and adjust its height as described above.



WARNING!

Important warning for the assistant: the push handle posts are safe to use only when the two levers **(1)** are tightly screwed, in fact, this condition prevents the posts from sliding up or down or even coming off accidentally.

Before pushing the wheelchair, make sure the two posts are safely locked.

Rehateam s.r.l. is not liable for any damage or injury caused by failure to heed this warning.

10.5 TIPPING AID

This accessory has been designed to enable the assistant to tip back the wheelchair without a great effort to facilitate negotiating small steps.

With one foot, press down on the plastic pedal (1) while, at the same time, applying downward force onto the push handles on the backrest (2).



IMPORTANT!

If you remove the plastic support of the tipping aid, you can mount the anti-tip wheel as explained in the chapter 10.3 *ANTI-TIP WHEEL*.



10.6 CRUTCH SUPPORT

This accessory holds crutches on the wheelchair without interfering with its normal use.

Put the ends of the crutches into the cup (1) and fasten their upper ends using the strap (2) fixed to the backrest post.

You can easily remove the post with the cup by pressing the pin (3) and sliding the post off the support (4).

To fit the post with the cup, insert the post into the support (4) until the pin (3) touches the support itself.

Then, press the pin and slide the post in until the pin locks in the support hole.



10.7 TRAY

This accessory can be assembled only if the wheelchair is equipped with armrests.

To install the tray, loosen bolt **(1)** of the support **(2)** on each side of the wheelchair until the hole is free to let the post go through.

Insert the tray posts **(3)** parallel into the supports and slide the tray in to the desired depth.

Finally, tighten the bolt **(1)** of both supports **(2)**.

To remove the tray, loosen the bolt **(1)** of the support **(2)** on each side of the wheelchair and slide the tray off.



WARNING!

Before using the tray, make sure it is firmly fixed.

Try to move it to-and-fro: if it does not move, it is ready for use.



10.8 PELVIC BELT AND HARNESS

Both the pelvic belt and the harness are accessories to this model and must therefore be ordered separately.

Both these belt systems are designed to assure the occupant greater safety and stability while seated in the chair.



IMPORTANT!

Both the pelvic belt and harness system are of particular use for occupants with little control of their upper body and who require extra support to be held securely in the seat during use.



DANGER!

Danger of serious injury or strangulation

A loose belt can allow the occupant to slide downward and pose a strangulation hazard.

The posture belt must be fitted by an experienced technician and the person responsible for prescribing its use must ensure its suitability.

The pelvic belt can be obtained from our network of authorised dealers.


PELVIC BELT

The pelvic belt **(1)** holds the occupant in the chair around the waist and leaves the upper body freedom of movement.

The pelvic belt is fastened to the wheelchair via two anchorage points **(2)** at the rear of the frame.

HARNESS

The harness **(3)** holds the occupant in the seat via two attachment points at waist level and two more at shoulder level.



The occupant's entire upper body is secured to the seat.

It is intended for use by occupants with particularly limited upper body mobility.

The harness is anchored to the wheelchair on four points at the rear frame **(2)** and at the backrest posts **(4)**.

USE

Using the pelvic belt and harness system is fairly simple.

Before sitting in the wheelchair, the occupant or assistant should open the buckles **(5)** by pressing on both sides of the locking buckle and pulling the two parts open.

Once seated in the wheelchair with the belt or harness in place, the occupant can close the buckles by fastening the two parts together.

Both belts are adjustable by pulling the flaps of the pelvic belt **(6)** or of the harness **(7)**.

It is also possible to adjust the length of the straps at the other buckles **(8)**.



WARNING!

Always check the correct functioning and tightness of the locking buckles before using the straps.

10.9 HEADREST

This accessory is designed to support the head.

It is suitable for folding wheelchairs; in fact, this type of headrest has a folding structure.

When you unfold the wheelchair, the headrest unfolds, too.

To put the headrest in its operating position **(1)**, pull it up until the two posts **(2)** are tilted downward, starting from the central block **(3)**.

Before folding the wheelchair **(4)**, you should push down the central block **(3)** of the headrest.

The occupant or the assistant can easily adjust the position of the headrest according to the occupant's need.

Unscrew the knob **(5)** to adjust the height.

Unscrew the knob **(6)** to adjust the depth.

Unscrew the knob **(7)** to adjust the pad by turning it up/down/right/left.



WARNING!

After each adjustment, screw the knob fast and make sure the system is stable

Always check the adjustment; in fact, if the system is unstable, it may be cause risky situations.



IMPORTANT!

After loosening each adjustment knob, you can also remove the corresponding parts.



10.10 LATERAL SUPPORT

This accessory is designed support the occupant's torso.

This lateral support is of the "swing-away" type so it can be swung back.

Turn the lever **(1)** by 90° and swing the lateral support to the opposite position where it will automatically lock.



WARNING!

Make sure the system is stable.

Always check the adjustment; in fact, if the system is unstable, it may be cause risky situations.



10.11 REAR WHEELS

10.11.1 General information

Depending on functional or aesthetic requirements, different rear wheels can be chosen in the order form and different types of hand rims and tyres can be fitted.



WARNING!

For each type of quick release wheel, always check that the quick release pin is inserted correctly. See chapter 5.5 *QUICK RELEASE REAR WHEELS*



*STANDARD
BASIC 24*



*STANDARD
PROGEO 22'*

10.11.2 Wheels Off-road

These wheels provide superior smoothness and comfort on rough or unpaved terrain.

They are also recommended if the wheelchair is equipped with a front propulsion device.

They are only supplied with their matching tyre and an aluminium hand rim.

You can choose this type of wheel as an additional kit (and they are interchangeable) to the main rear wheels you choose for your wheelchair.



10.11.3 Wheels Fat-wheel

These wheels enable the wheelchair to be used on more difficult terrain.

They are also recommended if the wheelchair is equipped with a front propulsion device.

They are only supplied with their matching tyre and an aluminium hand rim.

You can choose this type of wheel as an additional kit (and they are interchangeable) to the main rear wheels you choose for your wheelchair.



10.11.4 Wheels with single drive

This accessory enables the occupant to push and control the wheelchair using only one rear wheel. This wheel is provided with two independent hand rims: the larger one controls the wheel on the same side whereas the smaller drives the opposite wheel via the shaft **(1)**.

To fold the wheelchair, you must first remove the shaft **(1)**.

To do so, pull one of its two ends to collapse it (it is internally spring-loaded) until it can be detached from the wheel's axle and completely removed.

To fit the drive shaft, insert one of its two ends onto an axle of the wheel, then, pull the shaft to compress it as much as necessary to be able to insert the other end into the axle of the other wheel.





IMPORTANT!

Check the closure of the two knobs **(2)** regularly.

These should not have any play but, at the same time, they should enable the drive shaft to compress and extend.

This is the ideal condition for driving as well as for inserting and removing drive shaft **(1)**.

10.11.5 22" and 24" wheels equipped with drum brakes

This accessory enables the assistant to brake the wheelchair also when it is in motion.

It is very useful on slopes along which may be difficult for the wheelchair to negotiate.

To brake the wheelchair, just pull up the lever **(1)** that is fixed on each push handle.

The stronger the pull, the more braking force is applied to the wheel.

Releasing the levers **(1)** automatically stops the braking action.

Drum brakes can also act as parking brakes.

Press and hold the lever **(1)**, then push forward the toothed lever **(2)** on the underside of the brake lever and finally release the lever **(1)**.

Now the toothed lever **(2)** keeps lever **(1)** locked and the brake action is maintained.

To release the brake from this position, press the lever **(1)** just enough to be able to release the toothed lever **(2)** and release it.





WARNING!

For controlled braking it is necessary to apply the same pressure to both levers.

In fact, the two brakes are independent, so if you brake with only one lever, the wheelchair will turn to that same side.

Pay particular attention when going downhill.

10.11.6 12" wheels (300mm)

With these wheels, the wheelchair becomes shorter and narrower, but they are not as smooth as the 22" or 24" wheels.

The tyres can be solid **(1)** or air-filled **(2)**.

With these wheels, occupants cannot push themselves.

The parking brake is fixed in the rear side of the wheelchair and it is designed to be used by the assistant.

In the version with drum brake **(3)**, see the instructions in chapter 10.11.5 22" and 24" wheels equipped with drum brakes



10.12 HAND RIM

Special hand rims are available for this Progeo wheelchair model. Their special features ensure improved grip and therefore a more efficient pushing motion, especially for occupants with limited strength or mobility of the hands and fingers.

1. Polished aluminium
2. Aluminium - rubber-coated
3. Chrome-plated steel
4. Brushed titanium
5. Ultra Grepp — hard rubber
6. Max Grepp — soft rubber
7. Rudder
8. Silicone ring to fit over the aluminium hand rim
9. Top Friction - with hard rubber band around outer perimeter
10. Nova H aluminium
11. Nova D aluminium coated with hard rubber

NOTE: the figure of each hand rim shows the indicative drawing of the section



WARNING!

When assembling these hand rims, there is a risk of pinching or crushing the fingers between the hand rim itself and the wheel rim.

Always perform this operation with great care.



10.13 TYRES

Different types of tyres are available as listed below.

Compatibility with the rear wheel depends on the size of the wheel itself.

1. High pressure smooth tyre
Light and reliable, suitable for most environments.
2. High pressure profile tyre
Light and reliable, more suitable for rough surfaces.
3. Marathon Plus
With anti-puncture protection, suitable for most environments.
4. Solid tyre
Suitable for most environments, it does not need inflation but its ride is less smooth. It has a little less grip, especially over smooth and wet surfaces and it absorbs less vibration when rolling on rough surfaces.



10.14 SPOKE GUARDS

Their function is mainly cosmetic but they also prevent occupants from accidentally inserting their fingers into the rear wheel spokes.

1. Colour
2. Dragon
3. Black
4. Transparent



10.15 CASTORS

Various types and sizes of castors are available for this model, depending on the occupant's functional needs and/or aesthetic preferences.

In general, the larger the diameter and width of the wheel, the easier it will be to negotiate rough terrain and small obstacles such as stones, small height differences of the order of 1 cm.

Soft castors offer a more comfortable drive than hard ones.

The pneumatic castor is the most comfortable but must be inflated regularly.

Hard casters absorb less vibration when driving on rough surfaces.

1. 4" Soft — 100x32
2. 5" Soft — 125x27
3. 5" Tyre — 127x29
4. 6" Soft — 150x32
5. 6" tyre — 150x32
6. 7" Soft — 175x40



WARNING!

Before using the wheelchair, always check that the pivoting of the castors does not interfere with other wheelchair components.

Check periodically that the castors are properly secured.



10.16 VANITY FLAP (seat/backrest bonding canvas)

With this accessory you can affix the backrest upholstery to the seat canvas; it is useful for avoiding a possible backward sliding of a cushion that is not properly fixed to the seat.

The application of this vanity flap is quite simple; by means of the male Velcro straps present, simply attach it to the inside of the backrest upholstery and to the female Velcro seat canvas straps.



10.17 REMOVABLE UTILITY BAG

This accessory is very useful for stowing items to carry with you, both when using the wheelchair and in other situations.

In fact, the bag is easily removable thanks to the support fixed to the wheelchair.


Depending on the space available, the bag may face backwards **(1)** or forwards **(2)** and is normally fixed in the front part of the wheelchair.

To remove the bag, turn the knob **(3)** from the locking position **(4)** until the slotted guide **(5)** comes free.

The knob's lever **(3a)** can be in different positions, in fact there are various locking positions.

Once slotted guide **(5)** is free, pull the bag upwards to remove it **(6)**.

To attach the bag, insert the support **(7)** along the slotted guide **(5)** and finally turn the knob **(3)** to a locking position **(4)** as indicated above.



10.18 FRAME PROTECTION PAD

This accessory is very useful for protecting the frame from possible impact. The application of this padding is quite simple, in fact it is sufficient to wrap it around the frame and fasten it with the Velcro strips.



11 MAINTENANCE

The expected lifetime of the wheelchair, conservatively estimated, is five years.

Periodic inspection of the wheelchair is essential to guarantee maximum performance and a long life.

Regular, thorough checks and proper use of the wheelchair will extend its life by many years.

To clean the aluminium or carbon fibre parts (frame, hand rims, brakes etc.), the backrest and the seat, we recommend using only a soft, damp cloth.



WARNING!

Within 6 months from purchase, we recommend going to an authorised dealer for a complete check-up of your wheelchair.

Failure to check the wheelchair will immediately invalidate the warranty and release the manufacturer from any liability for possible damage.



WARNING!

We recommend paying particular attention to the cleaning of the hand rims of the rear wheels, which become dirty and dusty very easily because of their constant use and nearness to the ground.

Careful cleaning the hand rims ensures a firm grip and thus safer operation.



CAUTION!

When cleaning the wheelchair, do not use abrasive cleaning agents or degreasing substances which could cause damage.



WARNING!

Sand, sea water, water containing chlorine or other aggressive environments, may damage bearings, bolts, nuts or any parts subject to movement.

We recommend avoiding contact with the above-mentioned aggressive agents and, in case of exposure, performing a careful washing and cleaning.

The warranty does not cover oxidized parts and damage caused by careless use.



WARNING!

Firmly tighten all screws and replace lock nuts that are used frequently.

In fact, with frequent loosening and tightening they tend to lose their effectiveness.



WARNING!

Have a complete check of the wheelchair carried out (at least every three months) by qualified personnel authorised for maintenance on PROGEO products.

For any questions relating to the adjustment and maintenance of your Progeowheelchair, the experienced technical staff at Rehateam s.r.l. are at the complete disposal of all our clients.

You can contact us directly at the address below: Rehateam s.r.l., Vicolo Negrelli 5, 31038 Castagnole di Paese TV, Italia, tel. 0039 0422 484657, mail: info_it@permobil.com

11.1 REPLACEMENT OF WORN PARTS

With its use, any wheelchair will require, in addition to the routine maintenance, further “unscheduled” intervention due to the normal wear and tear of components.

Such servicing is closely linked to the amount of time for which it is used and how it is handled (e.g. use on uneven and rough terrain, in salty environments, etc.).

11.1.1 Repairing a puncture

In the event of a puncture, remove the wheel and Remove the tyre using bicycle tyre levers.

Remove the inner tube and repair it using a standard puncture repair kit and the same procedure used for repairing a normal bicycle inner tube.

If repair is not possible then you will need to replace the inner tube.

The rear wheel tyres should be replaced whenever excessive or irregular wear is noticed since this degrades the wheelchair's performance.

To re-fit an inner tube and tyre onto the wheel rim, it is necessary to partially inflate the inner tube. Insert the inflation valve into the hole provided on the rim.

Using both hands and bicycle tyre wrenches, insert the inner tube into the tyre and work the edge of the tyre over the wheel rim ensuring that all parts are fitted correctly.

While re-fitting the tyre, be careful not to pinch the inner tube.

Finally, inflate to the correct pressure.

In case of solid tyres, when worn out, they should be replaced with new ones.

11.1.2 Spare parts

For the replacement of parts due to wear and tear or breakage (or simply for the purchase of accessories) all the necessary spare parts to keep your wheelchair in perfect working order will remain readily available.



IMPORTANT!

All spare parts can be ordered through our authorised dealers. Adjustment may be necessary after replacement of parts.

11.2 CHECKING THE PARTS

As a daily check, we recommend the following operations:

- Check the tyre pressure.
- Check the quick release axles.

- Check the folding system (seat and/or backrest).
- Check the footplate.
- Check the parking brakes.
- General check of all screws.
- The condition of all parts subjected to wear and tear.

11.3 DISINFECTION AND REUSE OF THE WHEELCHAIR

When the wheelchair is used regularly, all surfaces that come into contact with the occupant should be treated frequently, or whenever the need arises, with a spray disinfectant.

It is not normally possible to guarantee the thorough disinfection of stitched seams.

We therefore recommend that the seat and backrest covers be changed and appropriately discarded in the event of bacterial contamination, in compliance with current regulations.

If the wheelchair is reused, the same type of treatment must be carried out very carefully before using the wheelchair again.



WARNING!

The first thing you should know is that cleaning is not the same as disinfecting.

By cleaning we remove germs and dirt from surfaces, but we do not kill germs, although removing them decreases the amount and risk of spreading infections.

On the other hand, with a disinfection, what we do is kill microorganisms such as bacteria through using chemicals.

This process does not necessarily clean dirty surfaces, but by killing germs on the surface, it can further reduce the risk of spreading the infection.

Keeping that in mind, the correct process to follow with the wheelchair is to first perform routine cleaning and then disinfect it using a sanitary surface disinfectant

These products are bactericidal and virucidal.

You can find more information about these products on the website of the Ministry of Health or other reliable sources

We also recommend that you carefully read the manufacturer's instructions for correct use and disinfection as well as to avoid possible contraindications.

11.4 EXTENDED STORAGE

If the wheelchair is not used for a medium-long period (over 4 months), it should be cleaned and stored in a dry and covered place and inside a box (preferably that of the original packaging).

At the time of re-use, it is necessary to carry out a general check of the wheelchair following the same points in paragraph 11.2 *CHECKING THE PARTS*

Before the occupant can use the wheelchair, carry out driving tests of the wheelchair to check for defects.

In case of malfunctions, faults or broken parts, consult an authorised dealer for the necessary repair.

11.5 DISPOSAL/RECYCLING OF PARTS



When the wheelchair is to be disposed of (end of life of the product), it is necessary to consider any prevailing local regulations for waste disposal or recycling.

This includes cleaning or disinfecting the wheelchair before disposal.

A description of the materials used is provided below.

- Aluminium: frame, forks, wheels, backrest tubes, footplate tubes
- Titanium: frame, backrest tubes
- Steel: bolts and nuts, quick release axle
- Plastic: footplates, side guards
- Rubber/PU: grips, tyres, wheels
- Carbon fibre: frame, side guards, footplates
- Upholstery: textile with expanded foam
- Upholstery: textile with expanded foam.

Packaging: plastic bags made of soft polyethylene, cardboard
Alternatively, you can return the backrest to your dealer for disposal.



WARNING!

In the event of disposal, be aware of any broken or damaged parts which could create a situation of potential danger with cutting or pinching injuries.

The use of suitable protection is recommended.

12 TROUBLESHOOTING GUIDE

With constant and prolonged use of the wheelchair, or after the adjustment of certain parts, some “defects” may emerge which can easily be eliminated by qualified personnel, or, where indicated by the ■ symbol, also by the occupant or assistant. In any case, we advise you to consult an authorised Rehateam s.r.l. dealer.

Problem	Cause of the problem	Solution	
The wheelchair does not go straight.	The front forks are not perpendicular to the ground.	Check castor angle	
	The castors are not set to the same height.	Check castor height.	
	Incorrect or unbalanced tyre pressure.	Inflate the tyres at the same and correct pressure	■
	The spokes are broken or loose.	Change the damaged spokes or tighten the loose ones.	
	The castor bearings are dirty or damaged.	Clean the bearings. Change the bearings.	■
The wheelchair tips up easily.	The rear wheel trim is set too far forward.	Adjust the rear wheel trim further rearward.	
	The wheelchair is excessively inclined.	Increase the rear height or decrease the front height.	
The parking brakes do not work properly.	Incorrect tyre pressure.	Check tyre pressure.	■
	Bad adjustment of brake position.	Check brake adjustment.	
The wheelchair fails to roll smoothly.	Incorrect tyre pressure.	Check tyre pressure.	■
	Worn out tyres.	Clean and lubricate the parts.	■
The wheelchair is difficult to unfold.	Over-tightened backrest bands.	Slacken the backrest bands.	■

13 TECHNICAL SPECIFICATIONS

- **SEAT WIDTH** 360; 390; 420; 450; 480; 510; 540 mm
- **SEAT DEPTH** 400; 425; 450 mm (*adjustable*)
- **FOOTREST DISTANCE** from 300 to 520 mm (*adjustable*)
- **BACKREST HEIGHT** from 270 to 470 mm (*adjustable*)
- **BACKREST ANGLE** standard frame: 84° 90° 96° (*adjustable*)
- **CAMBER** 0°
- **FRONT HEIGHT** from 400 to 530 mm (*adjustable*)
- **REAR HEIGHT** from 380 to 530 mm (*adjustable*)
- **TRIM (centre of gravity)** 10; 30; 50 mm (*adjustable*)
- **FRONT FRAME ANGLE** 105°
- **TOTAL WIDTH** seat width + 200 mm
- **TOTAL LENGTH (standard configuration)** approx. 970 mm
- **WHEELCHAIR WEIGHT (lightest configuration)** from approx. 15.8 kg
- **WEIGHT WITHOUT EXTRACTABLE PARTS (lightest configuration)** from approx. 9.0 kg (*without rear wheels, armrests, footrests*)
- **WEIGHT PAIR OF STANDARD REAR WHEELS** approx. 3.5 kg
- **WEIGHT PAIR OF EXTRACTABLE FOOTRESTS** approx. 1.5 kg
- **WEIGHT PAIR OF ARMRESTS** from approx. 1.5 to 2.4 kg
- **MAXIMUM LOAD** 125 kg
- **MINIMUM OVERALL DIMENSIONS (wheelchair folded)** 300 mm

14 WARRANTY

The warranty agreement exists only between Rehateam s.r.l. and its authorised dealers. For this reason, the client may not make warranty claims directly. The following warranty conditions are therefore presented solely for information purposes.

General conditions of the warranty

Rehateam s.r.l. provides assistance on its products provided they have been used correctly and that adequate maintenance has been carried out on all parts of the wheelchair. The warranty covers all material and production defects provided that such defects can be shown to have been caused prior to delivery of the product to the authorised dealer.

How to make use of warranty rights

To benefit from all warranty rights on all our products, the authorised distributor/retailer must carry out an inspection, within 7 days from the date of delivery, on the products received, in order to ascertain any manufacturing defects, and secondly, if a manufacturing defect is found, the authorised distributor/retailer must immediately report it in writing to Rehateam s.r.l.

All defects which, in spite of scrupulous control, have been identified only after the expiry of the period indicated above, must also be reported in writing to Rehateam s.r.l.

Warranty period

On this model of wheelchair Rehateam s.r.l. provides a guarantee of 2 years on the frame, and 2 years on all other components and accessories, starting from the date of delivery, excluding those components that are subject to normal wear and tear during everyday use.

Repair of defects and replacement

The warranty on defects on contact parts is at the complete discretion of Rehateam s.r.l., either for the repair of the defect or for the replacement of the part itself. The authorised distributor in cases of simple repairs may take action independently to eliminate the defect or bring the defect to the attention of Rehateam s.r.l. in specific cases. With reference to our production line that is labelled with the trademarks Progeo and Physio, since these products are manufactured to occupant specifications and with artisanal and not industrial procedures, minor imperfections, minor colour and carbon fibre weave discrepancies are not to be considered as defects.

Limits of the warranty

The warranty does not cover additional costs (e.g. repair, packing, labour costs, incidental costs).

The following are not covered by the warranty:

- Damage caused during shipment and not notified to the carrier at the moment of delivery.
- Repairs carried out by unauthorised dealers or personnel.
- Parts subject to wear - consult your distributor/dealer for a list of these parts.
- Damage to property or injury to persons caused during use of our products.
- Damage caused maliciously or where the buyer is at fault or resulting from incorrect or improper use of the product.
- Damage caused to the wheelchair, to objects and to persons resulting from any device or object, which is mounted on or added to the wheelchair without written authorisation from Rehateam s.r.l.

All claims for compensation are excluded from the warranty except for those specifically mentioned in the preceding paragraphs of this chapter.

Rehateam s.r.l. accepts no responsibility for failure to respect or execute the indications established in the individual contracts, if the following events have prevented and/or made it impossible to respect the terms of the contract itself: embargoes, import and export bans on contracted products, legal regulations, strikes and lock-outs, shortages of raw materials, accidents or other force majeure circumstances.

Rehateam s.r.l. is not obliged to disclose any technical variations made to its products, which may be subject to modifications and updates as deemed necessary.

15 WARRANTY CERTIFICATE



REHATEAM s.r.l.



Vicolo Negrelli 5 — 31038 Castagnole di Paese (TV) — Italy



0039 0422 484657



0039 0422 484661



info_it@permobil.com



www.rehateamprogeo.com

DISTRIBUTOR / DEALER

Stamp

Date of delivery to user:

MEDICAL DEVICE
CLASS I



product label

permobil

eng-UK



www.permobil.com