



Yoga
User manual | British English

Some photographs and/or pictures in this manual may show different products from the reference product.

However, all instructions are applicable, regardless of the product differences.

Total or partial reproduction of this manual is forbidden without written authorisation from Rehateam s.r.l.

Produced and published by Permobil

Edition: 0.1

Date: 2024-12-17

Order no: eng-UK

1	INTRODUCTION	5			
2	FLAME-RETARDANT PROPERTIES, SYMBOLS	6			
2.1	FLAMMABILITY	6	5.7.5	Single flip-up aluminium footrest	23
2.2	SYMBOLS USED	6	5.7.6	Single flip-up carbon fibre footrest	23
3	PRODUCT LABELS	7	5.7.7	Single plastic footrest with automatic folding	23
4	INTENDED PURPOSES AND CONDITIONS OF USE	8	5.7.8	Single aluminium platform with automatic folding	24
5	PUTTING THE WHEELCHAIR INTO SERVICE AND REGULAR USE	10	5.8	CALF STRAP	24
5.1	GENERAL INFORMATION	10	5.9	ADJUSTABLE BACKREST UPHOLSTERY	25
5.2	TYRE PRESSURE	11	5.10	ARMRESTS	25
5.3	BRAKES	12	5.10.1	General information	25
5.4	TYPE OF BRAKES	14	5.10.2	Tubular armrests with steel support	27
5.4.1	Curved lever push to lock brake	14	5.11	CLOTHING SIDE GUARD	29
5.4.2	Straight lever push to lock brake	14	5.12	WHEELCHAIR PARTS FASTENING AND LIFETIME	29
5.4.3	Straight lever pull to lock brake	15	5.12.1	Parts fastening	29
5.4.4	Extended lever push to lock or pull to lock brake	15	5.12.2	Wheelchair lifetime	30
5.5	QUICK RELEASE REAR WHEELS	16	6	SAFETY	31
5.6	FOLDING AND UNFOLDING THE WHEELCHAIR	17	6.1	PREVENT ACCIDENTS	32
5.6.1	Unfolding	18	6.1.1	Movements	32
5.6.2	Folding	19	6.1.2	Wheelchair speed	32
5.7	FOOTREST	20	6.1.3	Brakes use	32
5.7.1	General information	20	6.2	MOVEMENTS: GETTING IN AND OUT OF THE WHEELCHAIR	33
5.7.2	Dual footrest	22	6.3	STARTING TO DRIVE YOUR WHEELCHAIR	35
5.7.3	Front frame brace band	22	6.4	OVERCOMING OBSTACLES: STEPS AND GRADIENTS	36
5.7.4	Single flip-up plastic footrest	22	6.4.1	Going up and down a step by yourself	37
			6.4.2	Negotiating a step with a carer	38
			6.5	NEGOTIATING SLOPES OR INCLINES	39
			6.6	ENSURING STABILITY	40

7	TRASPORT OF THE WHEELCHAIR	43
7.1	TRASPORT OF THE OCCUPIED WHEELCHAIR	44
7.2	TRANSPORT WITH A CARER	44
7.3	LOADING THE WHEELCHAIR WITHOUT ASSISTANCE	45
8	DESCRIPTION OF PARTS	46
9	ADJUSTMENT	47
9.1	GENERAL INFORMATION	47
9.2	POSSIBLE ADJUSTMENTS	48
10	ACCESSORIES	49
10.1	GENERAL INFORMATION	49
10.2	NARROW PASSAGE TRANSIT WHEELS	49
10.3	ANTI-TIP WHEEL (with steel support)	50
10.4	CARER'S PUSH HANDLEBAR	52
10.5	TIPPING AID	53
10.6	CRUTCH SUPPORT	53
10.7	FOLDING PUSH-HANDLES	54
10.8	QUICK RELEASE FORKS	54
10.9	TETRACLIP	55
10.10	PELVIC BELT AND HARNESS	56
10.11	LUMBAR SUPPORT	58
10.12	LATERAL SUPPORT	58
10.13	REAR WHEELS	59
10.13.1	General information	59
10.13.2	Wheels Spinergy Flexrim	59
10.13.3	Wheels Off-road	59
10.13.4	Wheels Fat-wheel	60
10.14	HAND RIM	60
10.15	TYRES	61

10.16	SPOKE GUARDS	62
10.17	FRONT CASTORS	62
10.18	VANITY FLAP (seat/backrest bonding canvas)	63
10.19	REMOVABLE UTILITY BAG	64
10.20	FRAME PROTECTION PAD	64

11	MAINTENANCE	65
11.1	REPLACEMENT OF WORN PARTS	66
11.1.1	Repairing a puncture	66
11.1.2	Spare parts	66
11.2	CHECK OF PARTS	66
11.3	DISINFECTION AND REUSE OF THE WHEELCHAIR	67
11.4	EXTENDED STORAGE	67
11.5	DISPOSAL/RECYCLING OF PARTS	68
12	TROUBLESHOOTING GUIDE	69
13	TECHNICAL SPECIFICATIONS	70
14	WARRANTY	71
15	WARRANTY CERTIFICATE	73

1 INTRODUCTION

At Rehateam, constant research into quality and creativity are the cornerstone of our business.

Rehateam s.r.l. is a serious, reliable company that is remain genuinely focused on providing total customer satisfaction.

We have become industry leaders by making excellence and service our top priorities.

We have become industry leaders by making excellence and service our top priorities, building all our models from carefully selected materials and providing multiple configurations to enable the utmost personalisation according to specific user needs.

We perform continuous, meticulous quality control and our testing procedures offer the highest possible quality combined with fast, reliable service.

We acknowledge that we owe our success to our customers and all who have believed in us and helped us establish that quality is the supreme differentiator.

FOR THE DEALER

This manual **MUST** be provided to the occupant of this wheelchair.

BEFORE giving the wheelchair to the occupant, the dealer **MUST** carry out a general check of all its fittings and check its functionality.

After this check, the dealer should complete the **WARRANTY CERTIFICATE** by applying the stamp of the company or represented organisation together with name, signature, and date.

This certificate must be retained by the occupant.

This product is intended for disabled occupants with intact visual, cognitive and reading abilities.

If some or all of these abilities are limited or impaired, the presence of an assistant person with appropriate abilities is always required.

FOR THE OCCUPANT and/or CARER

BEFORE using this wheelchair, please **READ** all parts of this manual and keep it for future reference.

Check that the warranty certificate has been fully compiled by the dealer and keep it with care.

NOTE

This document is also available in PDF format for the visually impaired at www.rehateamprogeo.com (scan the QR code here aside).



2 FLAME-RETARDANT PROPERTIES, SYMBOLS

2.1 FLAMMABILITY

This product has been tested in accordance with the standard EN12183. This standard also includes a flammability test.

The fire reaction of upholstered parts complies with classes: M2 and M3.

2.2 SYMBOLS USED



DANGER!

Indicates a potentially hazardous situation, which, if not avoided, can result in death or serious personal injury.



CAUTION!

This symbol indicates an action which is forbidden or not recommended.



WARNING!

Indicates a potentially hazardous situation, which, if not avoided, may cause minor personal injuries or property damage.
















IMPORTANT!

This symbol indicates important instructions or special information.

3 PRODUCT LABELS



 Progeo brand logo	 "UDI" code (Unique Device Identification)
 "MD" mark (Medical Device)	 Year/month of production
 "CE" mark	 Serial number
XXXXXXXXXX Product name	 Ambient usage temperature
 Crash test approved Suitable for transporting the occupied wheelchair in a vehicle	 Max Maximum load
 Not suitable for transporting the occupied wheelchair in a vehicle	 Max Maximum slope in use
 Read the user's manual	 Manufacturer data

4 INTENDED PURPOSES AND CONDITIONS OF USE

This model is a self-propelled wheelchair which means that it **can be moved manually** by driving the hand rims on the rear wheels or by a carer using the push handles.

If self-propulsion is via the lower limbs, take care that the user's feet and/or legs do not collide with any part of the wheelchair.

It may be used for both rehabilitative and active purposes and is therefore suitable for daily use by both occupants with severe disabilities (with limited movement of the legs, arms or torso) and "active" occupants (capable of carrying out activities independently).

It is used to assure the independent mobility or assisted movement of an occupant with reduced or no use of one or more parts of the body.

Given the ample range of possible accessories and configurations, the wheelchair can be used safely both indoors (flat, gym, school, library, etc.) and outdoors (street, pavement, yard, etc.) with ambient temperatures between -30 and +60 °C.

It is possible that the surfaces of the wheelchair, which may come into direct contact with the skin of the occupant/carers, will exceed a surface temperature of 41°C (e.g. exposure to the sun), so take care before use.

If the occupant is affected by particularly severe disabilities or unable to move autonomously, the help of an assistant is always advisable.

We also discourage use over hilly, particularly soft, sandy or uneven ground, on slopes exceeding the recommended gradient and in acidic environments.



WARNING!

During daily use of the wheelchair, the occupant may come across situations such as rough or irregular terrains (gravel, cobblestones, holes on the ground, etc.) or slippery terrain (wet, sandy, dusty, oily hard surfaces, etc.).

In all such situations and, in general, in all cases where use of the wheelchair may be difficult or risky (for the wheelchair and its occupant), it is always advisable to move carefully and smoothly without sudden acceleration or hard braking.

When the difficulty or risk is high, the help of an assistant is always advisable.

The maximum slope that can be negotiated to prevent the wheelchair from tipping over backwards is 3°C (6%).

All footrests require the occupant to wear socks or shoes, so do not rest bare feet on them.

The seating surface, in cloth or other material if available for this model, is not designed to be directly sat on by the occupant; it is therefore necessary to provide a cushion for the occupant to sit on and which must be ordered separately according to the occupant's specific needs.



WARNING!

Crushing, holes and clearances: The wheelchair meets the requirements of safety distances. If certain functional requirements do not comply with these distances, be careful when using it (toe and foot traps).



IMPORTANT!

In certain configurations, the wheelchair may exceed **700 mm in width**.

In such cases, under certain circumstances, it may not be possible to use some or all of the available escape routes from a building.

If the wheelchair configuration causes the wheelchair width to exceed 700 mm, it may be difficult or impossible to travel on public transport.

In certain configurations, the wheelchair may exceed **1200 mm in width**.

Therefore, it may be difficult, if not impossible, to use public transport and/or transport the occupied wheelchair (if provided).

Given its dimensional and structural characteristics, the wheelchair model is ideal for adolescents and adults. occupants

Maximum weight load for this model is: 125 kg.

5 PUTTING THE WHEELCHAIR INTO SERVICE AND REGULAR USE

5.1 GENERAL INFORMATION



WARNING!

It is important that all the wheelchair's functional parts are CHECKED ONLY BY QUALIFIED AND AUTHORISED PERSONNEL, both to assess its suitability and to provide the correct instructions for use to occupant and carer where applicable.

If any part of the wheelchair is found to be defective, the occupant or assistant must consult the dealer to solve the problem.

Non-observance of the indications/instructions in this manual may lead to malfunctions and/or damage to the wheelchair and/or the occupant and/or third parties.

is not liable for any damage to the wheelchair and/or the occupant and/or third parties caused by or resulting from the non-observance of the indications/instructions.



CAUTION!

Use of the wheelchair requires some of its parts to be handled.

Instructions for the correct handling of the wheelchair's components are provided in the corresponding sections of this document.

Please follow these instructions to avoid injuries.



CAUTION!

The installation and use of electrification, mechanical and/or electrical drive devices (lifts, stairlift) or any other accessory not provided by Rehateam s.r.l. at the time of delivery may result in stresses that were not foreseen during product testing and therefore the integrity and complete safety of the product is not guaranteed.

In case of installation of any additional accessory to the product delivered by Rehateam s.r.l., the latter must be guaranteed by your dealer, the accessory device manufacturer, and approved by Rehateam s.r.l..

5.2 TYRE PRESSURE

The tyres must always be inflated to the right pressure to ensure that the wheelchair brakes and rolls properly.

Use your fingers to remove the plastic cap which covers the inflation valve on the rim of the rear wheel.

Use a compressor or pump equipped with a pressure gauge to bring the pressure up to the correct level.

Screw the plastic cap back on.




WARNING!

Check for correct tyre pressure on a weekly basis as indicated on the tyre (an intermediate value between min. and max. is recommended).

Proper tyre pressure makes the wheelchair considerably smoother and more controllable to manoeuvre.

Inadequate tyre pressure (especially if lower than prescribed) degrades the effectiveness of the parking brakes (the wheelchair may not remain stationary).





CAUTION!

Do not exceed the pressure indicated on the tyre because you may damage the tyre and the inner tube.

Always respect what indicated on the tyre.



IMPORTANT!

The front castors may also be of the "pneumatic" type; in this case, follow the same instructions.

Solid tyres do not require any inflation.

This type of tyre requires less maintenance and reduces puncture repair or replacement costs.

Compared to tubed or tubular tyres, they are not as smooth during normal use and absorb less vibration when rolling on uneven ground.

5.3 BRAKES



CAUTION!

The parking brakes are a safety device and their efficiency must be checked daily.

The pressure brake acts directly on the tyre by means of a lever, which is why braking efficiency depends on correct tyre inflation pressure.

We recommend that you check your tyre pressure regularly to ensure that it is always correct.

The maximum safe gradient for the parking brakes to work properly is 3° (6%).

The parking brake is a safety device, and you must not use it while the wheelchair is moving as this could cause the wheelchair to tip over with possible injury to the occupant.



IMPORTANT!

To check the effectiveness of the brakes, apply them as described below depending on the type of brake, and, with the person sitting in the wheelchair, try to gently push the wheelchair.

If the rear wheels do not turn or if they turn with great difficulty and still prevent unintentional movement of the wheelchair, the parking brakes are acting as required.

Check that both brakes, once applied, do not release accidentally.

Applying and releasing the parking brake may require more or less force depending on the type of brake chosen.

Occupants with reduced upper limb strength should use the extended lever.



WARNING!

Take care not to pinch/crush your fingers when applying and releasing the brakes, but also in any other situation where your hands come close to the brakes.

5.4 TYPE OF BRAKES

5.4.1 Curved lever push to lock brake

To engage the brake, push the lever forward until it locks.

To release the brake, pull the lever backward.

The brake automatically returns to its resting position.



5.4.2 Straight lever push to lock brake

To engage the brake, push the lever forward until it locks.

To release the brake, pull the lever backward.

The brake automatically returns to its resting position.




5.4.3 Straight lever pull to lock brake

To engage the brake, pull the lever backward until it locks.

To release the brake, push the lever frontward.

The brake automatically returns to its resting position.



5.4.4 Extended lever push to lock or pull to lock brake

Push to lock brake: To engage the brake, push the lever forward until it locks.

To release the brake, pull the lever backward.

The brake automatically returns to its resting position.

Pull to lock brake: To engage the brake, pull the lever backward until it locks.


To release the brake, push the lever frontward.

The brake automatically returns to its resting position.

Lever extension: The lever may be folded to prevent the extended lever from obstructing the occupant's wheelchair transfer manoeuvre.

To do so, pull the lever first up and then push it forward.

To return it to the normal using position, turn the lever up and then push it down.



5.5 QUICK RELEASE REAR WHEELS

The rear quick release wheels enable you to reduce the wheelchair's bulk and save space, thus facilitating storage in the car or, if the wheelchair is equipped with the "narrow passage wheels" accessory, to reduce its width and length in order to access narrow spaces.

Check that the quick release pins on the rear wheels are correctly engaged without anyone sitting on the wheelchair.

The parking brakes must be released.

Raise the rear wheel by lifting the wheelchair slightly off the ground and grasping the backrest post handle.

Grasp the rear wheel hub with your hand (without pressing down on the release pin), taking care to avoid inserting your fingers between the spokes.

Push and pull the wheel (as indicated by the arrow) to check that it fits properly: the wheel must not come off.

If the wheel comes off during this check, the axle must be adjusted.

Adjustment should only be carried out by qualified and authorised personnel.




WARNING!

Checking the quick release axles of the rear wheels is one of the procedures that the occupant should perform periodically.

If the quick release axles are not correctly inserted and/or adjusted, they could cause damage to components and the wheel could accidentally come off, causing the wheelchair to tip over with possibly injury to the occupant.

To remove the wheel, first release the parking brakes.

Lift the wheelchair wheel slightly off the ground, using one hand to grasp the backrest knob to facilitate the operation.



Grasp the wheel by the spokes around the hub, press the release pin button with your thumb and, without releasing, extract the wheel fully.

To fit the wheel, first release the brake.

Slightly lift the wheelchair from the ground on one side by the backrest knob.

With your hand, take hold of the rear wheel's hub and, with your thumb, press the button of the quick release pin, keep it pressed and push the pin fully into its seat.

Release the axle quick release pin button (a click indicates correct engagement).



WARNING!

Always release the brake before removing and inserting the rear wheel. Always check the correct engagement of the quick release axles. The holes for the rear wheel positioning may pinch or crush your fingers. Take care whenever your hands are close to such holes.

5.6 FOLDING AND UNFOLDING THE WHEELCHAIR

This model may be folded by acting on the backrest posts only.



CAUTION!

Before using the wheelchair, always check the seat is properly locked in the open position.

While folding/unfolding the backrest, take care not to crush your fingers between the backrest post and the side guard.

5.6.1 Unfolding

(1) Grasp the backrest handle grip with one hand and the front frame with the other hand.

Turn the backrest backward until the wheelchair is completely open: you should hear a click.

Pull the other backrest post backward until you hear the click also on this side.

To check that the wheelchair has opened correctly, grasp the front frame with one hand on the front frame and the other at a backrest handle and try to move the backrest forwards and backwards.

If the backrest post is correctly attached, there is only a small amount of play in the backrest post.

Check both backrest posts and that for both side plates, the quick release axle **(4)** on the inner side itself is correctly inserted.

To make the operation easier, you should lift one rear wheel to reduce ground friction. To do so, take hold on the backrest and slightly lift the wheelchair.

If there is a split or single folding footrest, lower the footrest completely (this operation is not necessary if a single folding footrest is fitted).



5.6.2 Folding

Remove the cushion or any other soft or rigid supports from the seat and the backrest.

Raise the footrest (this operation is not necessary if the wheelchair is equipped with an automatic closure footrest).

(1) Grasp the cord located at the bottom of the backrest posts and pull it up or backward.

(2) With the other hand, take hold of the backrest handle grip and push it forward a little while keeping the cord pulled upward.

(3) Release the cord and, with one hand, pull the seat fabric up and, with the other, fold the backrest post down until the wheelchair has completely folded

(4).



WARNING!

If the wheelchair is fitted with a dual footrest or flip-up footrest, raise it fully before folding the wheelchair (this operation is not necessary if the wheelchair is equipped with an automatic folding footrest).

If you neglect this procedure, you may cause damage to the frame and/or other parts of the wheelchair. Rehateam s.r.l. shall not be liable for any damage caused by failure to properly execute this procedure.



IMPORTANT!

If the wheelchair is equipped with tubular armrests, they must be folded behind the backrest or removed.



5.7 FOOTREST

5.7.1 General information



WARNING!

When transferring to and from the wheelchair, never keep your feet on the footrest; excessive weight on the footrest may result in it breaking with harm to the occupant.

After each operation, check that the footrest is always in the correct operating position.

Dual footrests are less resistant than the single version and they are therefore not suitable for occupants subject to strong spasms and/or contractions.



WARNING!

Do not use the wheelchair barefoot to avoid possible abrasions and/or burns to the feet.

Check the good condition of the platform surfaces as, in the event of contact with other surfaces (ground, walls, furniture, etc.), potentially dangerous sharp areas could be created.

Depending on the occupant's needs, the wheelchair can be provided with either of the above footrest versions.

In the order form, according to the type of footrest, you can choose its position (internal, external or intermediate).

Single footrests are more rigid than the dual versions and they are recommended for occupants subject to frequent spasms or who apply a lot of force to the footrest.

All footrests, except those with automatic closure, when raised, enable the occupants to rest their feet on the ground more easily, especially when transferring to and from the wheelchair.

Flipping up the footrest is a very easy operation you can perform with one hand by pulling the footrest plate upward.

To put the footrest back to the operating position, just push it down and, for a single footrest, make sure it locks in the support arm as indicated for each type of footrest.


Other information on footrest

To comply with the rule UNI EN 12183 paragraph 6.1, when dual footrests are fitted, it is necessary to maintain the correct clearance between the two footrest plates when lowered:

The footrest clearance "**C**" must not exceed 35 mm for wheelchairs used by adults **(1)**.

The footrest clearance "**C**" must not exceed 25 mm for wheelchairs used by children **(2)**.

Here image on the side shows the correct dual footrest size "**B**" with respect to seat width "**A**":



5.7.2 Dual footrest

Make sure both footrests are fully lowered before using the wheelchair.



5.7.3 Front frame brace band

If the wheelchair is equipped with a dual footrest, a stretched calf band is fixed to the fork support to prevent the front frames from separating when the user is transferring into or from it.



WARNING!

Never use the stretched calf band as support for your feet and never remove it.


Without it, in fact, the front frame can separate so much as to alter the wheelchair's geometry, even under normal movement.

Rehateam s.r.l. is not responsible for malfunction or damage to the wheelchair or the user if the wheelchair is used without the stretched calf band.



5.7.4 Single flip-up plastic footrest

Make sure that the locking pin (1) is properly inserted in the recess of the support (2).



5.7.5 Single flip-up aluminium footrest

Make sure that the locking pin (1) is properly inserted in the slot of support (2).



5.7.6 Single flip-up carbon fibre footrest

Make sure that the locking pin (1) is properly inserted in the slot of support (2).

The carbon fibre footrest can be swung 180°.

To perform such operation, just raise the plate out of its locking support and turn it to the opposite position.

Now you can either fold the wheelchair or lock the footrest in its support.



5.7.7 Single plastic footrest with automatic folding

This check is not necessary because the footrest opens automatically together with the seat.



5.7.8 Single aluminium platform with automatic folding

This check is not necessary because the footrest opens automatically together with the seat.



5.8 CALF STRAP


In some cases, depending on the diameter of the front wheel and the width of the wheelchair, the front wheel may come into contact with the heel of the occupant's foot when turning.

To prevent this contact and prevent the wheelchair from tipping over, the wheelchair is equipped with a heel strap that must be adjusted so that the occupant's feet never come into contact with the front wheels.



IMPORTANT!

You can easily adjust the calf strap using the provided Velcro straps.



5.9 ADJUSTABLE BACKREST UPHOLSTERY

The backrest can be adjusted to the occupant requirements by tensioning or slackening the two special Velcro straps fitted inside the backrest itself.

To carry out the adjustment, lift or remove the upholstery, tighten, or slacken the Velcro straps as required and replace the upholstery.



CAUTION!

The combination of the inclination of the seat with the inclination of the backrest posts and the inclination of a postural backrest with respect to the posts, could result in an angle greater than 25° with respect to the vertical.

In that case, the wheelchair must have a headrest or arrangement to secure it as intended for this model.



IMPORTANT!

If the wheelchair is provided with a postural backrest, follow the instructions of the postural system.

5.10 ARMRESTS

5.10.1 General information

This model can be equipped with armrests.

The armrests are particularly useful accessories for people with reduced hand or arm mobility and for occupants with limited upper body movement who require better stability once seated in the wheelchair.





CAUTION!

Do not use the armrest as a hand hold to raise the wheelchair; this may cause the armrests to come off from their supports, damaging the wheelchair and harming the occupant.

Do not use the armrest for support when transferring to or from the wheelchair.

On the contrary, we recommend removing the armrest on the transfer side to facilitate the manoeuvre.



WARNING!

Armrest pads may contain phthalates that are potentially hazardous to health if ingested.

Avoid scratching the padding and putting any residue on your fingers into your mouth or, worse still, nibbling or sucking on the padding itself (as may occur with children).



WARNING!

Always pay attention to the positioning of hands and arms on the armrest to avoid accidents.

5.10.2 Tubular armrests with steel support

You can swing this type of armrest away on the side and remove it to make transfer easier.

Clothing guards are with this type of armrest.

POSITION OF USE

To check that the armrest is in the correct operating position shown as **(1)**, take hold of the armrest with one hand **(2)** and try to move it left and right.

If it is properly secured, the armrest will not turn but will have just a little play.

The armrest is correctly secured when the two flat faces of the support **(3)** locate between the two walls of the support **(4)**.

The correctly attached steel support does not allow unintentional removal of the armrest.

Check by pulling the armrest upwards. The armrest must stop after approximately 1 cm.

ROTATION

To facilitate wheelchair transfers, you can swing by raising it by approximately 1 cm and swinging it away (part **(4)** of the support moves away from part **(3)**) but it will not come off).

If you wish, you can swing the armrest away until it locks in its opposite position (part **(4)** locks into part **(3)**).

To return the armrest to its position of use, repeat the above steps.

During rotation, the two supports **(3)** and **(4)** remain raised by about 1 cm, but do not detach from each other.

REMOVAL/INSERTION

To remove the armrest, with one finger, move the pin **(8)** to align with the hole and then pull the armrest up.

To insert the armrest; insert the pin **(7)** of the support into the hole of part **(3)** and turn the armrest until it locks as described above.





WARNING!

Always check that the armrests are secured when they are in their operating position **(1)** (facing forwards and parallel with the sides of the wheelchair).

In all other positions (except when locked in their rear-facing position), they are free to turn and for this reason, you should always ensure that they are correctly positioned before using the wheelchair.

Failure to follow this instruction may cause the wheelchair or the occupant to overbalance.



IMPORTANT!

The supports are precisely machined so that inserting and removing the armrest is possible only if the movement is perfectly vertical.

5.11 CLOTHING SIDE GUARD

The clothing guards can be in plastic or, much more resistant, in carbon fibre. They can be the “mudguard” type **(1)** or “straight” **(2)**.

Should the lateral clothing guard be of the type “mudguard”, these are originally fixed with its edge at approximately 5 mm above the tyre **(3)** to prevent the occupant from inserting his/her fingers between tyre and mudguard.

To remove the extractable lateral clothing guard **(4)**, just pull it of vertically.

To mount it **(5)**, align the two slots **(6)** on the lateral clothing guard to the bolts **(7)** and slide it all the way down the slots.



CAUTION!

Check the good condition of the surface of the clothing guard as they could be sharp and cause damage to the user.



IMPORTANT!

Do not use the clothing guard for support when transferring to or from the wheelchair.

5.12 WHEELCHAIR PARTS FASTENING AND LIFETIME

5.12.1 Parts fastening

When delivered to the occupant, the wheelchair is checked by the retailer who thus ensures it is in perfect condition.



However, daily use of the wheelchair may cause bolts and nuts to loosen.



WARNING!

We recommend checking that all hardware is correctly tightened at least once a month. If necessary, consult an authorised Progeo dealer or technician for maintenance assistance.

At each maintenance cycle, ensure that all nuts, bolts, and screws are correctly tightened.

Failure to observe this warning could result in damage to the occupant, third parties or the vehicle, for which Rehateam s.r.l. shall not be liable.

5.12.2 Wheelchair lifetime

Progeo wheelchairs are used on a daily basis and they are consequently subject to continuous stress that causes inevitable wear and tear of the parts. With this in mind, Progeo wheelchairs have a lifetime of 5 years provided they receive thorough periodic maintenance.

The lifetime will be considerably extended if the wheelchair is used only indoors or not on daily basis.

6 SAFETY



CAUTION!

The Progeo wheelchair is a medical device and therefore it **MUST NOT BE LENT NOR GIVEN TO OTHER USERS**, even for short period.

It is forbidden to make modifications to the wheelchair without our approval.

Modifications to the wheelchair may create problems of stability and resistance with possible harm to the occupant.



WARNING!

Rehateam s.r.l. will not be liable for failure to comply with the instructions or recommendations set out in this manual and any such activity will result in the immediate cancellation of the manufacturer's warranty.

BURN HAZARD! Wheelchair components may heat up if they are exposed to strong sunlight.



IMPORTANT!

Notify the manufacturer in the event of serious accidents in relation to the device.

Manufacturer: Rehateam s.r.l., Vicolo Negrelli 5, 31038 Castagnole di Paese TV, Italia, tel. 0039 0422 484657, mail: info_it@permobil.com

6.1 PREVENT ACCIDENTS

6.1.1 Movements

Avoid sudden, jerky movements when using the wheelchair as this could cause it to overbalance.

If your path is obstructed by obstacles, avoid sudden movements (such as abrupt braking). There is increased risk of skidding on irregular or damp ground.

Passing over obstacles as steps and ramps should be done carefully (ask a carer for help).

For greater safety, a carer should help to push the wheelchair when moving across inclined or potentially dangerous terrain.

6.1.2 Wheelchair speed

Always adjust your speed in relation to the type of terrain and conditions.

As general rule, we advise a constant, regular speed avoiding sudden accelerations or changes in direction.

6.1.3 Brakes use

The parking brakes have been designed to hold the wheelchair stable when it is stationary.



CAUTION!

In the interest of safety, you should not use the parking brakes while the wheelchair is in motion.



WARNING!

When using the brakes, great care should be taken to avoid injury to the hands or fingers.

6.2 MOVEMENTS: GETTING IN AND OUT OF THE WHEELCHAIR

Wheelchair transfers must always be performed with extreme care and caution, even by experienced occupants, and only after receiving instruction from fully trained specialists. If wheelchair transfers are difficult, an assistant should be lend help. All these actions should in any case be performed taking account of the degree of your disability and your consequent level of autonomy.

The general rules to respect while getting in or out of the wheelchair safely are:

Ensure the wheelchair is parked on a solid, flat, or regular surface.

Do not get in or out of the wheelchair while on hills or uneven ground which could render the wheelchair unstable and cause the occupant to fall and/or the wheelchair to overbalance. The parking brakes should be engaged.

Move your feet off the footrest to get down. Move your feet close to the footrest to go up.

If the wheelchair is equipped with flip-up footrest (single or dual), raise it to facilitate the transfer.

- (1)** Lean on the wheelchair and, where possible, on a stable object nearby. Use the force of your arms to raise and move your body.
- (2)** If the occupant is not able to perform this movement alone, or if the movement is to be performed on an uneven surface, a carer should lend help. Carers who help with wheelchair transfers must always observe the utmost care.
- Such operations must be performed only after receiving instruction from fully trained specialists.



6.3 STARTING TO DRIVE YOUR WHEELCHAIR

FINDING THE CENTRE OF GRAVITY

Your ability to control and safely tip the wheelchair mainly depends on your seating position and the location of the centre of gravity with respect to the rear wheels.

The correct position of the rear wheels depends on different factors such as the occupant's weight, type of disability and the occupant's driving skills.

The manufacturer delivers the wheelchair and the position of the rear wheels (configuration) as requested in the order form.


To find the point at which the chair will tip back easily and, consequently, to find the safe manoeuvring range, proceed as described below.

Starting from your seated wheelchair position:

Move forward by firmly grasping the rear wheel rims.

Then move the hand rims slightly backwards and, while moving your weight backwards, push the hand rims forward with a jerk to make the front of the wheelchair rise

At this point, by moving your body gently backwards and forwards, while moving the rear wheel backwards and forwards using the hand rims, you will be able to locate the centre of gravity and the different points at which you can achieve controlled tipping and risk dangerous overbalance of the wheelchair.



WARNING!

Finding the point at which the chair will tip back safely requires the presence of an assistant who must stand behind the wheelchair ready to prevent it from overbalancing.



WARNING!

If you hang a backpack, a bag or any other object on the push handlebars or any rear area of the wheelchair, the resulting uneven weight distribution can significantly affect the risk of overbalancing.

TO PREVENT THE WHEELCHAIR FROM OVERBALANCING DURING NORMAL USE, WE RECOMMEND FITTING AN ANTI-TIPPING SYSTEM.

6.4 OVERCOMING OBSTACLES: STEPS AND GRADIENTS



WARNING!

Always take the utmost care when dealing with the situations described below.

Before negotiating a step, or overcoming an obstacle in general, on your own (i.e. without the presence of a carer), you must have good control of your wheelchair and know how to Perform controlled tipping (wheelies), keeping your balance on the rear wheels.

Before tackling a step, or more generally overcoming an obstacle, make sure the anti-tipping system is in the neutral position.



DANGER!

Never attempt to negotiate steps or other obstacles that are over 20 cm in height.

When faced with particularly challenging obstacles, always seek assistance to ensure that such manoeuvres are performed with ease and safely.

6.4.1 Going up and down a step by yourself

GOING DOWN


Take the wheelchair right up to the obstacle so that the front wheels are as close as possible to it.

Raise the front castors by tilting the wheelchair while maintaining balance.

Slide the rear wheels very carefully over the edge.

Always keep your hands firmly on the hand rims during this step.

Once the rear wheels have descended the obstacle, tilt the wheelchair forward until the front castors touch the ground again.



GOING UP

Bring the wheelchair right up to the leading edge of the obstacle.

Raise the front castors and tip the chair back while maintaining balance.

Using the hand rims, carefully move the rear wheels forwards until they touch the edge of the step.

Tip the wheelchair forward until the front castors rest on the upper surface of the obstacle.

Lean forward in the chair so that the rear of the chair is lighter.

Grasp the hand rims tightly and move the rear wheels forwards that will cause the rear wheel to slide over the edge of the step lifting the rear of the wheelchair off the ground resulting in the overcoming of the obstacle.



6.4.2 Negotiating a step with a carer

GOING DOWN

Take the wheelchair right up to the obstacle so that the front wheels are as close as possible to it.


Grip the rear handles tightly and push down to raise the front of the wheelchair.

Keep the wheelchair in this position and accompany it down the step.

The occupant seated in the chair can help the assistant by using the hand rims.

The occupant seated in the chair can help the assistant by using the hand rims.

At the end of the descent, tip the wheelchair forward so that the front castors touch the ground again.




GOING UP

Move backwards towards the step so that the rear wheels of the wheelchair touch it.

Grip the rear handles tightly and push down to raise the front of the wheelchair.

Pull hard while keeping the wheelchair tilted (with the front wheels raised to prevent the occupant from slipping out of the chair) until the rear wheels are over the step.

Keeping the wheelchair tilted, move it away from the step enough to enable the front castors to touch the same level ground.



GOING UP AND DOWN A SERIES OF STEPS



DANGER!

In the case of stairs with more than one step, two assistants are always required.


The first assistant stands behind the wheelchair and holds the push handles.

The second assistant grasps a fixed part of the front frame, guiding the wheelchair manoeuvre from the front.



IMPORTANT!

To make raising the front of wheelchair easier for the assistant, the wheelchair should be fitted with the “tipping aid” accessory.



6.5 NEGOTIATING SLOPES OR INCLINES

GOING DOWNHILL

To go downhill safely, the occupant should keep constant, controlled speed, and direction over the entire slope.

Approach the slope at a moderate speed.

Keep your weight backwards in the chair to avoid slipping out.

Hold on to the hand rims on the rear wheels and allow them to slide slowly through your hands.

The occupant should be able to stop the wheelchair at any time simply by blocking the motion of the hand rims.



GOING UPHILL

Driving up any kind of slope should be done with extreme care.

The occupant must move at speed by generating force on the rear wheel while at the same time maintaining control over the wheelchair.

Lean forwards and keep your weight forward in the chair to prevent the wheelchair from tipping over backwards.

Grasp the hand rims tightly.

Drive up the ramp by pushing the wheelchair quickly and forcefully on both hand rims.

Do not jerk, but move steadily and as smoothly as possible.

Do not swing your upper body.



IMPORTANT!

We recommend that occupants who have not yet completely mastered safe use of the wheelchair should have the anti-tip system fitted, which is useful for preventing the wheelchair from tipping over backwards.



CAUTION!

The maximum gradient that can be attempted in complete safety is 3° (6%).

6.6 ENSURING STABILITY

You will encounter situations which will require you to lean out of the wheelchair.

These apparently simple movements could, if not performed with care, result in a loss of stability and possible tipping over of the wheelchair.



To achieve maximum occupant control of the wheelchair, we illustrate some situations which may emerge during normal daily use of the wheelchair. Please pay particular attention to the following to maintain balance and stability at all times.



WARNING!

Do not lean too far forward; you may fall out of the wheelchair.

Do not move forward by sliding your pelvis across the seat cushion to reach objects that are too far away: you could fall off the wheelchair.

Do not lean too far back over the backrest, the wheelchair could overbalance.

Do not lean too far out to the side, the wheelchair could overbalance.

LEANING FORWARDS


Ensure that the front castors are pointing forwards.

To do this, move the wheelchair forwards and then backwards.

Apply the parking brakes to prevent the chair from moving suddenly during the movement.

Moving the weight of the body excessively forward may cause the wheelchair to tip up onto the front castors and tipping over possibly causing damage to the chair and harming the occupant.

To ensure greater stability, the occupant should hold on to the wheelchair with his or her free hand. Do not apply the parking brakes.



LEANING BACKWARDS

Ensure that the front castors are pointing forwards.

To do this, move the wheelchair forwards and then backwards.

Do not apply the parking brakes.


Lean backwards without changing your sitting position.



LYING SIDeways

Do not move your upper body beyond the rear wheel as the wheelchair could overbalance.

For greater safety and stability, hold on to the wheelchair with your free hand.



7 TRASPORT OF THE WHEELCHAIR

There is no single best way to load the wheelchair into a vehicle.

Factors include the type and degree of the occupant's disability (his/her ability to control his/her upper body, arm, and hand movements), physical strength (the operation may be too difficult for an older person or a child) and the type of vehicle that will be used.

It should be clear that all these factors are too many to indicate a single precise procedure to fit all cases.

For this reason, we will limit ourselves to general advice.



WARNING!

Perform all car loading activities with extreme care and only after receiving instruction from specialist service personnel at our authorised dealers.

If these procedures seem to be unsafe or difficult, we recommend asking a carer for help.



CAUTION!

Never transport an occupant sitting on the wheelchair in a vehicle unless the wheelchair is certified for the purpose according to the requirements of the standard.

If the occupant must be transported while seated on the wheelchair, remember that the wheelchairs are not supplied with seatbelts. Any safety belts for vehicle transport must be installed only by companies that are specialised in vehicle transformations.

7.1 TRASPORT OF THE OCCUPIED WHEELCHAIR

Always make sure the wheelchair is suitable to transporting the user in a vehicle.

The symbols on the product label indicate if the wheelchair is or not crash tested and, accordingly, if it can be used for the transport of the user on a motor vehicle.

WARNING! Only wheelchairs marked with this symbol are judged suitable for use with the passenger on board the wheelchair.



7.2 TRANSPORT WITH A CARER

In cases where the physical limitations of the occupant prevent him/her from moving and loading the wheelchair into the car on his/her own, the help of an assistant will be required.

The lightness and foldability of this model are very important because they enable the wheelchair to be moved easily even by persons with limited physical strength.

Remove all removable parts.

Release the brakes and remove the rear wheels.

Fold the wheelchair (seat and/or backrest).

Take hold of the chair with two hands on the two fixed parts of the frame to balance the weight of the wheelchair and make it easier to lift.

Loading the wheelchair without assistance.

The following actions are intended for people with good control of their upper body, arms, and hands as well as sufficient strength to perform the movements in complete safety.

Thanks to its reduced encumbrance, the wheelchair can be loaded in the boot of the car or between the front and rear seats.



images are indicative only

In larger means of transport, you can load the wheelchair without reducing its encumbrance.

7.3 LOADING THE WHEELCHAIR WITHOUT ASSISTANCE

The following actions are intended for people with good control of their upper body, arms, and hands as well as sufficient strength to perform the movements in complete safety. Remove all removable parts.

Open the door on the side where you will get into the car from (e.g. the driver's door), then move the chair as close as possible to the seat.

Engage the brakes and get into the car.

Once you are inside the car, release the wheelchair brakes and pull out the rear wheels.

Fold the wheelchair (seat and/or backrest).

Take hold of the chair with two hands on the two fixed parts of the frame to balance the weight of the wheelchair and make it easier to lift.



Loading the wheelchair without assistance.

The following actions are intended for people with good control of their upper body, arms, and hands as well as sufficient strength to perform the movements in complete safety.



images are indicative only

8 DESCRIPTION OF PARTS



1. Backrest post with handle
2. Backrest upholstery + Velcro straps for adjustment
3. Tyre
4. Hand rim
5. Parking brake
6. Side plate
7. Backrest support
8. Quick release pin
9. Rear wheel
10. Rear wheel plate
11. Fork support
12. Fork
13. Clothing side guard
14. Seat canvas
15. Crossbar
16. Front frame
17. Calf strap
18. Footrest post
19. Footrest
20. Castor
21. Front frame brace band
22. Front frame brace band

9 ADJUSTMENT

9.1 GENERAL INFORMATION

The Progeo wheelchair is a medical device manufactured according to a patient-specific data sheet compiled by qualified personnel.

Rehateam s.r.l. manufactures, tests and delivers the wheelchair to the dealer with the measurements and characteristics required in the order data sheet.



CAUTION!

We strongly advise occupants against lending the wheelchair to other users even for brief periods.

Lending your wheelchair to others could cause it to function unsafely and tip over with potentially serious consequences both for the wheelchair and its occupant.

The wheelchair has a configuration and dimensions that are specially designed for the occupant who purchases it, and the characteristics of stability, manoeuvrability and resistance are guaranteed only for that occupant.

It is forbidden to carry out any modifications to the original design, even if they are possible.



CAUTION!

Adjustments may be carried out exclusively by qualified and authorised by Rehateam s.r.l. personnel.

Any adjustments and/or modifications carried out by non-authorised persons will immediately void the product's warranty and relieve Rehateam s.r.l. from any liability concerning any malfunction and/or damage due to such adjustments/modifications.

Always contact Rehateam s.r.l. and its technicians for any non-standard requirements or modifications to enable them to assess such modifications and verify that they will not compromise the normal and safe use of the wheelchair.

Any modification of the original parameters and configuration could seriously compromise the safe operation of the wheelchair causing damage the wheelchair itself and harm to its occupant.

After every adjustment made to the wheelchair, check carefully that all parts are correctly fixed. Check that all screws and nuts are tightened and that all moving parts are functioning correctly.

After any adjustment, always test the wheelchair before normal use, possibly assisted by a carer or a technician.

The possible adjustments for this model are listed below.

The instructions to properly carry out the adjustments are included in the corresponding service manual that is available on the web site www.rehateamprogeo.com

9.2 POSSIBLE ADJUSTMENTS

- Rear height
- Front height
- Trim (centre of gravity)
- Fork angle
- Backrest height
- Backrest inclination
- Footrest distance
- Footrest position
- Footrest inclination
- Brakes
- Convergence

10 ACCESSORIES

10.1 GENERAL INFORMATION

Each model can be equipped with various accessories that should be evaluated according to the needs of the occupant and assistant.

Some of the images in this manual may show a wheelchair model different from the reference one.

However, all instructions are applicable, regardless of the model differences.

10.2 NARROW PASSAGE TRANSIT WHEELS

This accessory becomes necessary when the dimensions of the wheelchair impede getting into or passing through narrow rooms such as a door or a lift.


To use this accessory, you should remove the rear wheels.

This makes the wheelchair narrower and shorter.

You can remove the transit wheel from its support. Press the pin **(1)** that holds it and slide the post off the support **(2)**.

To put on the transit wheel, insert the post in the support **(2)** until the pin **(1)** touches the support itself.

Then, press the pin and slide the post in until the pin locks in the support hole.





WARNING!

For a correct assembly of the transit wheel, remember that when inserting it into the support, the pin **(1)** should always face toward the inside of the wheelchair.

Remember that while using the transit wheels (with the rear wheels removed), the parking brakes will not work at all.

The stability of the wheelchair resting on the wheels for narrow passages is reduced, therefore pay great attention and seek help from a carer.

10.3 ANTI-TIP WHEEL (with steel support)

This accessory is designed to prevent the wheelchair from tipping backwards.

To be efficient, it is fixed at approximately 2-3 cm from the ground.

You can have the anti-tip wheel can be fitted on one or, for more greater safety, on both sides of the wheelchair.

POSITION OF USE

To check for the correct operating position **(1)**, take hold of the post with one hand and try to move it right and left: the post should not turn (there may be a little play).

Then, push the post downward: the post goes down by approximately 1 cm, but it must not come off.

ROTATION

To swing the anti-tip wheel **(2)**, push it down by approximately 1 cm (the part **(3)** of the support moves away from the part **(4)** without coming off) and turn it sideways.

When you turn the anti-tip wheel by 180° (opposite position **(8)**), the support automatically locks (the part **(3)** locks in the part **(4)**).



To return the anti-tipping device to its position of use, repeat the same steps as described above.

REMOVAL/INSERTION

To remove the anti-tip wheel, push the pin **(6)** that you find in the upper side of the support **(6)** push the post down.

To put the anti-tip wheel on, insert the part **(7)** of the support in the hole of the part **(4)** and turn the post until the two parts locks together.

NEUTRAL POSITION

Set the anti-tipping device to the neutral position (8) if there is a need, e.g. to climb steps.



WARNING!

Verify the efficiency of the system with an assistant.

The front of the wheelchair must be raised until the anti-tip wheel touches the ground and prevents the wheelchair from tipping.

Take great care when performing this operation.

Make sure it is correctly fitted (in the use position) before using the wheelchair.



IMPORTANT!

When overcoming obstacles or steps, the anti-tipping device must be placed in the neutral position.

10.4 CARER'S PUSH HANDLEBAR

You can assemble these push handles only on a backrest posts that originally come without handles.

You can adjust their height to offer the assistant a better handhold and pushing position.

The assistant or the occupant can easily adjust and remove them.

To adjust their height, unscrew the lever **(1)**, slide the post up or down to the desired height, position the handle backwards and finally screw the lever **(1)** tightly.

The handle must not turn unintentionally.

To remove the push handle, unscrew the lever **(1)** and slide the post off the support.

To fit the push handle on the wheelchair, insert the post into the support and adjust its height as described above.




WARNING!

Important warning for the carer: the push handle posts are safe to use only when the two levers **(1)** are tightly screwed, in fact, this condition prevents the posts from sliding up or down or even coming off accidentally.

Before pushing the wheelchair, make sure the two posts are safely locked.

Rehateam s.r.l. is not liable for any damage or injury caused by failure to heed this warning.



10.5 TIPPING AID


This accessory has been designed to enable the assistant to tip back the wheelchair without a great effort to facilitate negotiating small steps.

With one foot, press down on the plastic support **(1)** while, at the same time, applying downward force onto the push handles on the backrest **(2)**.

You can remove the tipping aid by pressing the pin **(4)** that secures it and sliding the post from the support **(3)**.

To fit the tipping aid, slide the post in the support **(3)** until the pin **(4)** touches the support.

Then, press the pin and slide the post in until the pin locks in the support hole.



10.6 CRUTCH SUPPORT

This accessory holds crutches on the wheelchair without interfering with its normal use.

Put the ends of the crutches into the cup **(1)** and fasten their upper ends using the strap **(2)** fixed to the backrest post.

You can easily remove the post with the cup by pressing the pin **(3)** and sliding the post off the support **(4)**.

To fit the post with the cup, insert the post into the support **(4)** until the pin **(3)** touches the support itself.

Then, press the pin and slide the post in until the pin locks in the support hole.



10.7 FOLDING PUSH-HANDLES

When folded down, these push handles reduce the wheelchair's overall volume and reduce possible interference with the occupant's arms when self-propelling the wheelchair.

When in the operating position **(1)**, they act just like normal handles.

To fold the handle down, press pin **(2)**, just beneath the handle and swing the handle down **(3)**.


The pin remains pressed down **(2a)**.

To swing the handle up to the operating position, just swing it up until the pin **(2)** automatically locks.



WARNING!

When the carer moves the wheelchair with the occupant sitting on it, make sure the push handles are safely locked in their operating position.



10.8 QUICK RELEASE FORKS


This accessory enables the bulk of the wheelchair to be further reduced.

With a double pair of forks, this accessory can be useful if you want to change the size of the front castors very easily and quickly to tackle various types of terrain.

To remove the fork, take hold of it as shown in the image, press the pin **(1)** that is located on the lower side of the quick release axle **(2)** and slide the fork off.

To insert it, press and hold the pin **(1)** and let the quick release axle **(2)** to the stop through the hole of the bearing **(3)**.

Finally, release the pin.





CAUTION!

Ensure that the fork is firmly seated by trying to remove it by pulling it downwards.

The fork must not come off.

There may be a just little play.

10.9 TETRACLIP

This is a special quick release axle with a ring lever. It simplifies wheel removal for occupants with particular problems affecting their hands and fingers.

The quick release works by turning the ring lever 90°

When the wheel is insert in the bushing, the ring lever must be set as in figure **(1)**, i.e. turned by 90° with respect to the quick release axle.

To remove the wheel, turn the ring lever so that it aligns with the quick release axle as shown in figure **(2)**.

Then remove the wheel.

To engage the wheel, the ring lever must be aligned as in figure **(2)**.

Insert the wheel and turn the ring lever by 90° as in figure **(1)**.



WARNING!

Checking the quick release axles of the rear wheels is one of the procedures that the occupant should perform periodically.

If the pins are not properly fitted, the rear wheel may come off during use, resulting in the wheelchair tipping over and possible injury to the occupant.

See also chapter 5.5 *QUICK RELEASE REAR WHEELS*.

10.10 PELVIC BELT AND HARNESS

Both the pelvic belt and the harness are accessories to this model and must therefore be ordered separately.

Both these belt systems are designed to assure the occupant greater safety and stability while seated in the chair.



IMPORTANT!

Both the pelvic belt and harness system are of particular use for occupants with little control of their upper body and who require extra support to be held securely in the seat during use.



DANGER!

Danger of serious injury or strangulation.

A loose belt can allow the occupant to slide downward and pose a strangulation hazard.

The posture belt must be fitted by an experienced technician and the person responsible for prescribing its use must ensure its suitability.

The pelvic belt can be obtained from our network of authorised dealers.

PELVIC BELT

The pelvic belt **(1)** holds the occupant in the chair around the waist and leaves the upper body freedom of movement.

The pelvic belt is fastened to the wheelchair via two anchorage points **(2)** at the rear of the frame.

HARNESS



The harness **(3)** holds the occupant's torso to the wheelchair through two attachment points located at the waist level and two more at the shoulder level.

The occupant's entire upper body is secured to the seat.

It is intended for use by occupants with particularly limited upper body mobility.

The harness is fastened to the wheelchair with four anchorage points: two at the rear of the frame **(2)** and two on the backrest posts **(4)**.

USE

Using the pelvic belt and harness system is fairly simple.

Before sitting in the wheelchair, the occupant or carer should open the buckles by pressing on both sides of the locking buckle **(5)** and pulling the two parts open.

Once seated in the wheelchair with the belt or harness in place, the occupant can close the buckles by fastening the two parts together.

Both straps are adjustable by pulling the ends of the pelvic belt **(6)** or the harness **(7)**.

Furthermore, it is possible to adjust the length of the straps at the other buckles **(8)**.



WARNING!

Always check the correct functioning and tightness of the locking hooks before using the straps.

10.11 LUMBAR SUPPORT

This is a pad that offers extra support to the lumbar area.

To position the lumbar support, raise the backrest upholstery on its front side.

Using the Velcro fastener, attach the lumbar support either to the backrest upholstery or to the backrest bands at the required height.

Finally, return the backrest upholstery back to its normal position.



10.12 LATERAL SUPPORT

This accessory is designed support the occupant's torso.

This lateral support is of the “swing-away” type so it can be swung it back.

Turn the lever **(1)** by 90° and swing the lateral support to the opposite position where it will automatically lock.



WARNING!

Make sure the system is stable.

Always check the adjustment; in fact, if the system is unstable, it may be cause risky situations.



10.13 REAR WHEELS

10.13.1 General information

Depending on your functional and/or appearance needs, you can choose one of the different rear wheels available for this model and fit different types of hand rim and covers.



WARNING!

For each type of quick release wheel, always check that the quick release axle is inserted correctly. See chapter 5.5 *QUICK RELEASE REAR WHEELS*

10.13.2 Wheels Spinergy Flexrim

This type of wheel has an aluminium and flexible rubber hand rim that significantly increases hand grip.

Different types of covers can be fitted

10.13.3 Wheels Off-road

These wheels provide superior smoothness and comfort on rough or unpaved terrain.

They are also recommended if the wheelchair is equipped with a front propulsion device.

They are only supplied with their matching tyre and an aluminium hand rim.

You can choose this type of wheel as an additional kit (and are interchangeable) to the main rear wheels you choose for your wheelchair.



10.13.4 Wheels Fat-wheel

These wheels enable the wheelchair to be used on more difficult terrain.

They are also recommended if the wheelchair is equipped with a front propulsion device.

They are only supplied with their matching tyre and an aluminium hand rim.

You can choose this type of wheel as an additional kit (and are interchangeable) to the main rear wheels you choose for your wheelchair.



10.14 HAND RIM

Special hand rims are available for this Progeo wheelchair model. Their special features ensure improved grip and therefore a more efficient pushing motion, especially for occupants with limited strength or mobility of the hands and fingers.

1. Polished aluminium
2. Aluminium - rubberised paintwork
3. Chrome-plated steel
4. Brushed titanium
5. Ultra Grepp — hard rubber
6. Max Grepp — soft rubber
7. Rudder
8. Silicone ring for application above the aluminium hand rim
9. Top Friction — with hard rubber band around the outer perimeter
10. Nova H in aluminium
11. Nova D made of hard rubber-coated aluminium

NOTE: The approximate section of each hand rim is drawn in the in the figure.





WARNING!

When assembling these hand rims, there is a risk of pinching or crushing the fingers between the hand rim itself and the wheel rim.

Always perform this operation with great care.

10.15 TYRES

Different types of tyres, listed below, are available.

Compatibility with the rear wheel depends on the size of the wheel itself.

1. High pressure tyre
Light and reliable, suitable for most environments.
2. High pressure profile tyre
Light and reliable, more suitable for rough surfaces.
3. Marathon Plus
With anti-puncture protection, suitable for most environments.
4. Solid tyre
Suitable for most environments, it does not need inflation but its ride is less smooth. It has a little less grip, especially over smooth and wet surfaces and it absorbs less vibration when rolling on rough surfaces.



10.16 SPOKE GUARDS

Their function is mainly cosmetic but they also prevent occupants from accidentally inserting their fingers into the rear wheel spokes.

1. Colour
2. Dragon
3. Black
4. Transparent



10.17 FRONT CASTORS

Various types and sizes of castors are available for this model, depending on the occupant's functional needs and/or aesthetic preferences.

In general, the larger the diameter and width of the wheel, the easier it will be to negotiate rough terrain and small obstacles such as stones, small height differences of the order of 1 cm.

Soft castors offer a more comfortable drive than hard ones.

The pneumatic castor is the most comfortable but must be inflated regularly.

Hard casters absorb less vibration when driving on rough surfaces.

1. 3" Sport— 80x24
2. 3" Active Soft— 80x34
3. 4" Soft— 100x32
4. 4" Active Sport— 100x21
5. 4" Active Soft— 100x32
6. 5" Soft— 125x27
7. 5" Active Soft— 125x34
8. 5" air-filled — 127x26
9. 6" Soft— 150x32
10. 6" air-filled — 150x32





WARNING!


Before using the wheelchair, always check that the pivoting of the front castors does not interfere with other wheelchair components.

Periodically check that the front castors are properly secured.

10.18 VANITY FLAP (seat/backrest bonding canvas)

With this accessory you can affix the backrest upholstery to the seat canvas; it is useful for avoiding a possible backward sliding of a cushion that is not properly fixed to the seat.

The application of this vanity flap is quite simple; by means of the male Velcro straps present, simply attach it to the inside of the backrest upholstery and to the female Velcro seat canvas straps.



10.19 REMOVABLE UTILITY BAG

This accessory is very useful for stowing items to carry with you, both when using the wheelchair and in other situations.

In fact, the bag is easily removable thanks to the support fixed to the wheelchair.


Depending on the space available, the bag may face backwards **(1)** or forwards **(2)** and is normally fixed in the front part of the wheelchair.

To remove the bag, turn the knob **(3)** from the locking position **(4)** until the slotted guide **(5)** comes free.

The knob lever **(3a)** can be in different positions, in fact there are various locking positions.

Once slotted guide **(5)** is free, pull the bag upwards to remove it **(6)**.

To attach the bag, insert the support **(7)** along the slotted guide **(5)** and finally turn the knob **(3)** to a locking position **(4)** as indicated above.



10.20 FRAME PROTECTION PAD

This accessory is very useful for protecting the frame from possible impact.

The application of this padding is quite simple, in fact it is sufficient to wrap it around the frame and fasten it with the Velcro strips.



11 MAINTENANCE

The expected lifetime of the wheelchair, conservatively estimated, is five years.

Periodic inspection of the wheelchair is essential to guarantee maximum performance and a long life.

Regular, thorough checks and proper use of the wheelchair will extend its life by many years.

To clean the aluminium or carbon fibre parts (frame, hand rims, brakes etc.), the backrest and the seat, we recommend using only a soft, damp cloth.



WARNING!

Within 6 months from purchase, we recommend going to an authorised PROGEO dealer for a complete check-up of your wheelchair.

Failure to check the wheelchair will immediately invalidate the warranty and release the manufacturer from any liability for possible damage.



WARNING!

We recommend paying particular attention to the cleaning of the hand rims of the rear wheels, which become dirty very easily because of their constant use and nearness to the ground.

Careful cleaning the hand rims ensures a firm grip and thus safer operation.



CAUTION!

When cleaning the wheelchair, do not use abrasive cleaning agents or degreasing substances which could cause damage.



WARNING!

Sand, sea water, water containing chlorine or other aggressive environments, may damage bearings, bolts, nuts or any parts subject to movement.

We recommend avoiding contact with the above-mentioned aggressive agents and, in case of exposure, performing a careful washing and cleaning.

The warranty does not cover oxidized parts and damage caused by careless use.



WARNING!

Firmly tighten all screws and replace lock nuts that are used frequently.

In fact, with frequent loosening and tightening they tend to lose their effectiveness.



WARNING!

Have a complete check of the wheelchair carried out (at least every three months) by qualified personnel authorised for maintenance on PROGEO products.

For any questions relating to the adjustment and maintenance of your Progeo wheelchair, the experienced technical staff at Rehateam s.r.l. are at the complete disposal of all our clients

You can contact us directly at the address below: Rehateam s.r.l., Vicolo Negrelli 5, 31038 Castagnole di Paese TV, Italia, tel. 0039 0422 484657, mail: info_it@permobil.com

11.1 REPLACEMENT OF WORN PARTS

With its use, any wheelchair will require, in addition to the routine maintenance, further “unscheduled” intervention due to the normal wear and tear of components.

Such servicing is closely linked to the amount of time for which it is used and how it is handled (e.g. use on uneven and rough terrain, in salty environments, etc.).

11.1.1 Repairing a puncture

In the event of a puncture, remove the wheel and Remove the tyre using bicycle tyre levers.

Remove the inner tube and repair it using a standard puncture repair kit and the same procedure used for repairing a normal bicycle inner tube.

If repair is not possible then you will need to replace the inner tube.

The rear wheel tyres should be replaced whenever excessive or irregular wear is noticed since this degrades the wheelchair's performance.

To re-fit an inner tube and tyre onto the wheel rim, it is necessary to partially inflate the inner tube. Insert the inflation valve into the hole provided on the rim.

Next, insert the valve into the hole on the rim and, using both hands and bicycle tyre wrenches, insert the inner tube into the tyre and work the edge of the tyre over the wheel rim ensuring that all parts are fitted correctly.

While re-fitting the tyre, be careful not to pinch the inner tube. Finally, inflate to the correct pressure.

In case of solid tyres, when worn out, they should be replaced with new ones.

11.1.2 Spare parts

For the replacement of parts due to wear and tear or breakage (or simply for the purchase of accessories) all the necessary spare parts to keep your wheelchair in perfect working order will remain readily available.



IMPORTANT!

All spare parts can be ordered through our authorised dealers. Adjustment may be necessary after replacing the parts.

11.2 CHECK OF PARTS

As a daily check, we recommend the following operations:

- Check the tyre pressure.

- Check the quick release axles.
- Check the folding system (seat and/or backrest).
- Check the footplate.
- Check the parking brakes.
- General check of all screws.
- The condition of all parts subjected to wear and tear.

11.3 DISINFECTION AND REUSE OF THE WHEELCHAIR

When the wheelchair is used regularly, all surfaces that come into contact with the occupant should be treated frequently, or whenever the need arises, with a spray disinfectant.

It is not normally possible to guarantee the thorough disinfection of stitched seams.

We therefore recommend that the seat and backrest covers be changed and appropriately discarded in the event of bacterial contamination, in compliance with current regulations.

If the wheelchair is reused, the same type of treatment must be carried out very carefully before using the wheelchair again.



WARNING!

The first thing you should know is that cleaning is not the same as disinfecting.

By cleaning we remove germs and dirt from surfaces, but we do not kill germs, although removing them decreases the amount and risk of spreading infections.

On the other hand, with a disinfection, what we do is kill microorganisms such as bacteria through using chemicals.

This process does not necessarily clean dirty surfaces, but by killing germs on the surface, it can further reduce the risk of spreading the infection.

Keeping that in mind, the correct process to follow with the wheelchair is to first perform routine cleaning and then disinfect it using a sanitary surface disinfectant.

These products are bactericidal and virucidal.

You can find more information about these products on the website of the Ministry of Health or other reliable sources.

We also recommend that you carefully read the manufacturer's instructions for correct use and disinfection as well as to avoid possible contraindications.

11.4 EXTENDED STORAGE

If the wheelchair is not used for a medium-long period (over 4 months), it should be cleaned and stored in a dry and covered place and inside a box (preferably that of the original packaging).

At the time of re-use, it is necessary to carry out a general check of the wheelchair following the same points of paragraph 11.2 *CHECK OF PARTS*.

Before the occupant can use the wheelchair, carry out driving tests of the wheelchair to check for defects.

In case of malfunctions, faults or broken parts, consult an authorised dealer for the necessary repair.

11.5 DISPOSAL/RECYCLING OF PARTS



When the wheelchair is to be disposed of (end of life of the product), it is necessary to consider any prevailing local regulations for waste disposal or recycling.

This includes cleaning or disinfecting the wheelchair before disposal.

A description of the materials used is provided below.

- Aluminium: frame, forks, wheels, backrest tubes, footplate tubes
- Titanium: frame, backrest tubes
- Steel: bolts and nuts, quick release axle
- Plastic: footplates, side guards
- Rubber/PU: grips, tyres, wheels
- Carbon fibre: frame, side guards, footplates
- Upholstery: textile with expanded foam
- Upholstery: textile with expanded foam.

Packaging: plastic bags made of soft polyethylene, cardboard
Alternatively, you can return the backrest to your dealer for disposal.



WARNING!

In the event of disposal, be aware of any broken or damaged parts which could create a situation of potential danger with cutting or pinching injuries.

The use of suitable protection is recommended.

12 TROUBLESHOOTING GUIDE

With constant and prolonged use of the wheelchair, or after the adjustment of certain parts, some “defects” may emerge which can easily be eliminated by qualified personnel, or, where indicated by the ■ symbol, also by the occupant or carer. In any case, we suggest you should consult a Rehateam s.r.l. authorised dealer.

Problem	Cause of the problem	Solution	
The wheelchair does not go straight.	The front forks are not perpendicular to the ground.	Check castor angle	
	The castors are not set to the same height.	Check castor height.	
	Incorrect or unbalanced tyre pressure.	Inflate the tyres at the same and correct pressure	■
	The spokes are broken or loose.	Change the damaged spokes or tighten the loose ones.	
	The castor bearings are dirty or damaged.	Clean the bearings. Change the bearings.	■
The wheelchair tips up easily.	The rear wheel trim is set too far forward.	Adjust the rear wheel trim further rearward.	
	The wheelchair is excessively inclined.	Increase the rear height or decrease the front height.	
The parking brakes do not work properly.	Incorrect tyre pressure.	Check tyre pressure.	■
	Bad adjustment of brake position.	Check brake adjustment.	
The wheelchair fails to roll smoothly.	Incorrect tyre pressure.	Check tyre pressure.	■
	Friction on the locking pin.	Clean and lubricate the parts.	■
The wheelchair is difficult to unfold.	Over-tightened backrest bands.	Slacken the backrest bands.	■
	Friction on the folding mechanism crossbar.	Clean and lubricate the folding mechanism rails under the seat. EGO – EGO CARBON – EGO CUSTOM	■

13 TECHNICAL SPECIFICATIONS

- **SEAT WIDTH** 360; 390; 420; 450 mm
- **SEAT DEPTH** 350; 375; 400, 425, 450, 475 mm
- **FOOTREST DISTANCE** from 350 to 540 mm (*adjustable*)
- **BACKREST HEIGHT** from 245 to 470 mm (*adjustable*)
- **BACKREST ANGLE** 4 positions: 3>90°;1<90° (*adjustable*)
- **CAMBER** 0°; 2°
- **FRONT HEIGHT** from 430 to 530 mm (*adjustable*)
- **REAR HEIGHT** from 410 to 490 mm (*adjustable*)
- **TRIM (centre of gravity)** 75; 95; 115 mm (*adjustable*)
- **FRONT FRAME ANGLE** 90°; 100°
- **TOTAL WIDTH** seat width + 170 mm (*with camber 0°*)
- **TOTAL LENGTH (std. config.)** approx. 880 mm
- **WHEELCHAIR WEIGHT (lightest config.)** from approx. 12.2 kg
- **WEIGHT WITHOUT REMOVABLE PARTS (lightest config.)** from approx. 9.6 kg (*without rear wheels*)
- **STANDARD REAR CASTOR PAIR WEIGHT** approx. 3.2 kg
- **ARMREST PAIR WEIGHT** approx. 1.5 kg
- **MAXIMUM LOAD** 125 kg
- **MINIMUM SIZE (folded wheelchair)** 280 mm (*with camber 0°*)

14 WARRANTY

The warranty agreement exists only between Rehateam s.r.l. and its authorised dealers. For this reason, the client may not make warranty claims directly (therefore, warranty conditions are presented solely for information purposes).

General conditions of the warranty

Rehateam s.r.l. provides assistance on its products provided they have been used correctly and that adequate maintenance has been carried out on all parts of the wheelchair. The warranty covers all material and production defects provided that such defects can be shown to have been caused prior to delivery of the product to the authorised dealer.

How to avail of your rights under the warranty

To benefit from all warranty rights on all our products, the authorised distributor/retailer must carry out an inspection, within 7 days from the date of delivery, on the products received, in order to ascertain any manufacturing defects, and secondly, if a manufacturing defect is found, the authorised distributor/retailer must immediately report it in writing to Rehateam s.r.l.

All defects which, in spite of scrupulous control, have been identified only after the expiry of the period indicated above, must also be reported in writing to Rehateam s.r.l..

Warranty period

On this model of wheelchair Rehateam s.r.l. provides a guarantee on the frame, and 2 years on all other components and accessories, starting from the date of delivery, excluding those components that are subject to normal wear and tear during everyday use.

Repair of defects and replacement

The warranty on defects on contact parts is at the complete discretion of Rehateam s.r.l. either for the repair of the defect or for the replacement of the part itself. The authorised distributor in cases of simple repairs may take action independently to eliminate the defect or bring the defect to the attention of Rehateam s.r.l. in specific cases. With reference to our production line that is labelled with the trademarks Progeo and Physio, since these products are manufactured to occupant specifications and with artisanal and not industrial procedures, minor imperfections, minor colour and carbon fibre weave discrepancies are not to be considered as defects.

Limits of the warranty

The warranty does not cover additional costs (e.g. repair, packing, labour costs, incidental costs).

The following are not covered by the warranty:

- Damage caused during shipment and not notified to the carrier at the moment of delivery.
- Repairs carried out by unauthorised dealers or personnel.
- Parts subject to wear — consult your dealer / retailer for a list of these parts.
- Damage to property or injury to persons caused during use of our products.
- Damage caused maliciously or where the buyer is at fault or resulting from incorrect or improper use of the product.
- Damage caused to the wheelchair, to objects and to persons resulting from any device or object, which is mounted on or added to the wheelchair without written authorisation from Rehateam s.r.l..

All claims for compensation are excluded from the warranty except for those specifically mentioned in the preceding paragraphs of this chapter.

Rehateam s.r.l. accepts no responsibility for failure to respect or execute the indications established in the individual contracts, if the following events have prevented and/or made it impossible to respect the terms of the contract itself: embargoes, import and export bans on contracted products, legal regulations, strikes and lock-outs, shortages of raw materials, accidents or other force majeure circumstances.

Rehateam s.r.l. is not obliged to disclose any technical variations made to its products, which may be subject to modifications and updates as deemed necessary.

15 WARRANTY CERTIFICATE



REHATEAM s.r.l.



Vicolo Negrelli 5 — 31038 Castagnole di Paese (TV) — Italy



0039 0422 484657



0039 0422 484661



info_it@permobil.com



www.rehateamprogeo.com

DEALER / RETAILER

Stamp

Date of delivery to user:

MEDICAL DEVICE
CLASS I



product label

permobil

eng-UK



www.permobil.com